

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011022.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 11 01:16:54 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042787.D 1000 =PP042785.D 750 =PP042786.D
 250 =PP042788.D 50 =PP042789.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.653	2.509	2.581	2.657	2.477	2.575 E4	3.18
2) SA	Decachlorobiphenyl	1.714	1.568	1.638	1.688	2.024	1.726 E4	10.17
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	6.598	5.961	6.184	6.539	7.587	6.574 E2	9.49
17) L4	AR-1242-2	0.970	0.877	0.911	0.976	1.011	0.949 E3	5.71
18) L4	AR-1242-3	5.792	5.329	5.590	5.784	6.515	5.802 E2	7.60
19) L4	AR-1242-4	4.933	4.456	4.684	4.835	4.354	4.652 E2	5.27
20) L4	AR-1242-5	4.778	4.435	4.641	4.604	4.272	4.546 E2	4.31
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =PP042787.D 1000 =PP042785.D 750 =PP042786.D
 250 =PP042788.D 50 =PP042789.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.653	2.508	2.567	2.663	2.590	2.596 E4	2.45
2) SA	Decachlorobiphenyl	1.981	1.782	1.863	2.099	2.117	1.968 E4	7.41
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011022.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 11 01:16:54 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042787.D 1000 =PP042785.D 750 =PP042786.D
 250 =PP042788.D 50 =PP042789.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	8.352	7.439	7.789	8.508	9.051	8.228 E2	7.65
17) L4	AR-1242-2	1.152	1.044	1.089	1.179	1.208	1.134 E3	5.92
18) L4	AR-1242-3	6.580	5.946	6.212	6.740	8.166	6.729 E2	12.80
19) L4	AR-1242-4	6.443	5.552	5.868	6.505	7.890	6.452 E2	13.91
20) L4	AR-1242-5	7.430	6.698	7.011	7.771	7.857	7.353 E2	6.74
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range