

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011022.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 11 01:32:56 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042792.D 1000 =PP042790.D 750 =PP042791.D
 250 =PP042793.D 50 =PP042794.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.613	2.531	2.561	2.621	2.434	2.552	E4	2.97
2) SA	Decachlorobiphenyl	1.660	1.606	1.627	1.658	1.789	1.668	E4	4.28
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	4.766	4.555	4.658	4.787	5.208	4.795	E2	5.19
22) L5	AR-1248-2	6.530	6.078	6.358	6.607	6.902	6.495	E2	4.69
23) L5	AR-1248-3	8.002	7.377	7.580	7.993	8.210	7.832	E2	4.37
24) L5	AR-1248-4	8.136	7.492	7.717	8.176	8.408	7.986	E2	4.66
25) L5	AR-1248-5	7.797	7.358	7.480	7.902	8.029	7.713	E2	3.69
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP042792.D 1000 =PP042790.D 750 =PP042791.D
 250 =PP042793.D 50 =PP042794.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.707	2.600	2.649	2.749	2.386	2.618	E4	5.41
2) SA	Decachlorobiphenyl	1.980	1.790	1.867	2.088	2.100	1.965	E4	6.90
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011022.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 11 01:32:56 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042792.D 1000 =PP042790.D 750 =PP042791.D
 250 =PP042793.D 50 =PP042794.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	5.914	5.445	5.766	6.137	7.035	6.059 E2	9.91
22) L5	AR-1248-2	0.840	0.752	0.794	0.889	1.004	0.856 E3	11.37
23) L5	AR-1248-3	0.880	0.792	0.838	0.931	1.035	0.895 E3	10.47
24) L5	AR-1248-4	1.004	0.914	0.954	1.035	1.022	0.986 E3	5.15
25) L5	AR-1248-5	1.001	0.909	0.946	1.037	1.072	0.993 E3	6.67
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range