

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011022.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 11 01:47:54 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042797.D 1000 =PP042795.D 750 =PP042796.D
 250 =PP042798.D 50 =PP042799.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.664	2.509	2.574	2.668	2.365	2.556 E4	4.93
2) SA	Decachlorobiphenyl	1.674	1.569	1.624	1.664	1.567	1.620 E4	3.14
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	0.890	0.795	0.833	0.917	1.091	0.905 E3	12.61
27) L6	AR-1254-2	1.314	1.188	1.246	1.373	1.397	1.304 E3	6.67
28) L6	AR-1254-3	1.368	1.235	1.297	1.412	1.529	1.368 E3	8.22
29) L6	AR-1254-4	0.950	0.857	0.881	0.977	1.070	0.947 E3	8.95
30) L6	AR-1254-5	1.029	0.931	0.970	1.058	1.177	1.033 E3	9.17
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =PP042797.D 1000 =PP042795.D 750 =PP042796.D
 250 =PP042798.D 50 =PP042799.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.718	2.609	2.657	2.756	2.642	2.677 E4	2.23
2) SA	Decachlorobiphenyl	2.000	1.802	1.871	2.120	2.191	1.997 E4	8.17
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP011022.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 11 01:47:54 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP042797.D 1000 =PP042795.D 750 =PP042796.D
 250 =PP042798.D 50 =PP042799.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.513	1.360	1.412	1.601	1.710	1.519 E3	9.32
27) L6	AR-1254-2	1.324	1.179	1.216	1.397	1.553	1.334 E3	11.26
28) L6	AR-1254-3	2.094	1.889	1.964	2.212	2.504	2.133 E3	11.33
29) L6	AR-1254-4	1.251	1.125	1.165	1.307	1.365	1.243 E3	7.93
30) L6	AR-1254-5	1.702	1.550	1.601	1.765	1.810	1.686 E3	6.48
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range