

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011022\
 Data File : P0084099.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Jan 2022 13:26
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 14:35:35 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jan 10 11:26:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.044	4.037	1944026	405099	19.688	19.811
2) SA Decachlor...	11.247	9.444	1280840	380688	19.255	19.489
Target Compounds						
9) L2 AR-1221-2	5.397	4.395	36661	5053	39.071	25.831 #
28) L6 AR-1254-3	7.906	6.795	31216	5088	5.544	2.978 #
29) L6 AR-1254-4	8.204	7.040	23944	5141	5.843	4.766
30) L6 AR-1254-5	8.639	7.476	29126	6852	6.788	4.590 #
33) L7 AR-1260-3	0.000	7.293	0	10185	N.D.	7.772 #
45) L9 AR-1268-5	0.000	9.188	0	2922	N.D.	0.399 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011022\
Data File : P0084099.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Jan 2022 13:26
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 14:35:35 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010722.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jan 10 11:26:21 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 1944026
Conc: 19.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.037 min
Delta R.T.: 0.000 min
Response: 405099
Conc: 19.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.247 min
Delta R.T.: -0.012 min
Response: 1280840
Conc: 19.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.444 min
Delta R.T.: -0.006 min
Response: 380688
Conc: 19.49 ng/ml



#9 AR-1221-2


R.T.: 5.397 min
Delta R.T.: 0.004 min
Response: 36661
Conc: 39.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 4.395 min
Delta R.T.: -0.010 min
Response: 5053
Conc: 25.83 ng/ml




#28 AR-1254-3

R.T.: 7.906 min
Delta R.T.: -0.003 min
Response: 31216
Conc: 5.54 ng/ml



#28 AR-1254-3


R.T.: 6.795 min
Delta R.T.: -0.004 min
Response: 5088
Conc: 2.98 ng/ml



#29 AR-1254-4


R.T.: 8.204 min
Delta R.T.: -0.004 min
Response: 23944
Conc: 5.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 7.040 min
Delta R.T.: -0.002 min
Response: 5141
Conc: 4.77 ng/ml




#30 AR-1254-5

R.T.: 8.639 min
Delta R.T.: -0.004 min
Response: 29126
Conc: 6.79 ng/ml



#30 AR-1254-5


R.T.: 7.476 min
Delta R.T.: -0.003 min
Response: 6852
Conc: 4.59 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 8.727 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 7.293 min
Delta R.T.: -0.003 min
Response: 10185
Conc: 7.77 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.880 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.188 min
Delta R.T.: -0.002 min
Response: 2922
Conc: 0.40 ng/ml