

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011022\
 Data File : PP042785.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Jan 2022 14:56
 Operator : AJ\MA
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 01:08:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 11 01:02:42 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.880	4.049	2508703	2508484	97.203	97.203
2) SA Decachlor...	10.815	9.388	1567929	1782231	95.557	94.707
Target Compounds						
16) L4 AR-1242-1	6.184	5.315	596095	743901	949.269	942.191
17) L4 AR-1242-2	6.207	5.336	876710	1043670	949.419	950.540
18) L4 AR-1242-3	6.273	5.530	532926	594584	958.363	949.359
19) L4 AR-1242-4	6.378	5.625	445583	555155	949.182	925.644
20) L4 AR-1242-5	7.156	6.192	443527	669799	962.793	948.175



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011022\
Data File : PP042785.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Jan 2022 14:56
Operator : AJ\MA
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 01:08:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011022.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 11 01:02:42 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.880 min
Delta R.T.: 0.000 min
Response: 2508703
Conc: 97.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.049 min
Delta R.T.: 0.000 min
Response: 2508484
Conc: 97.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.815 min
Delta R.T.: 0.000 min
Response: 1567929
Conc: 95.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.388 min
Delta R.T.: 0.000 min
Response: 1782231
Conc: 94.71 ng/ml



#16 AR-1242-1


R.T.: 6.184 min
Delta R.T.: 0.000 min
Response: 596095
Conc: 949.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#16 AR-1242-1

R.T.: 5.315 min
Delta R.T.: 0.000 min
Response: 743901
Conc: 942.19 ng/ml




#17 AR-1242-2

R.T.: 6.207 min
Delta R.T.: 0.000 min
Response: 876710
Conc: 949.42 ng/ml



#17 AR-1242-2


R.T.: 5.336 min
Delta R.T.: 0.000 min
Response: 1043670
Conc: 950.54 ng/ml



#18 AR-1242-3


R.T.: 6.273 min
Delta R.T.: 0.000 min
Response: 532926
Conc: 958.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#18 AR-1242-3

R.T.: 5.530 min
Delta R.T.: 0.000 min
Response: 594584
Conc: 949.36 ng/ml




#19 AR-1242-4

R.T.: 6.378 min
Delta R.T.: 0.000 min
Response: 445583
Conc: 949.18 ng/ml



#19 AR-1242-4


R.T.: 5.625 min
Delta R.T.: 0.000 min
Response: 555155
Conc: 925.64 ng/ml



#20 AR-1242-5

R.T.: 7.156 min
Delta R.T.: 0.000 min
Response: 443527
Conc: 962.79 ng/ml

Instrument :
ECD_O
ClientSampleId :



#20 AR-1242-5

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 669799
Conc: 948.17 ng/ml