

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011022\
 Data File : PP042798.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Jan 2022 18:31
 Operator : AJ\MA
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 01:44:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011022.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 11 01:36:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.881	4.049	667109	689095	25.618	25.662
2) SA Decachlor...	10.815	9.387	416079	529879	25.483	27.201
Target Compounds						
26) L6 AR-1254-1	7.091	6.192	229311	400258	267.032	272.046
27) L6 AR-1254-2	7.319	6.351	343151	349295	268.055	273.095
28) L6 AR-1254-3	7.696	6.775	353038	552891	265.791	271.039
29) L6 AR-1254-4	7.990	7.015	244160	326697	266.538	269.540
30) L6 AR-1254-5	8.419	7.447	264502	441350	265.361	266.743



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP011022\
Data File : PP042798.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Jan 2022 18:31
Operator : AJ\MA
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 01:44:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP011022.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 11 01:36:44 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.881 min
Delta R.T.: 0.000 min
Response: 667109
Conc: 25.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.049 min
Delta R.T.: 0.000 min
Response: 689095
Conc: 25.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.815 min
Delta R.T.: 0.000 min
Response: 416079
Conc: 25.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.387 min
Delta R.T.: 0.000 min
Response: 529879
Conc: 27.20 ng/ml



#26 AR-1254-1


R.T.: 7.091 min
Delta R.T.: 0.000 min
Response: 229311
Conc: 267.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 6.192 min
Delta R.T.: 0.000 min
Response: 400258
Conc: 272.05 ng/ml




#27 AR-1254-2

R.T.: 7.319 min
Delta R.T.: 0.000 min
Response: 343151
Conc: 268.06 ng/ml



#27 AR-1254-2


R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 349295
Conc: 273.09 ng/ml



#28 AR-1254-3


R.T.: 7.696 min
Delta R.T.: 0.000 min
Response: 353038
Conc: 265.79 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.775 min
Delta R.T.: 0.000 min
Response: 552891
Conc: 271.04 ng/ml




#29 AR-1254-4

R.T.: 7.990 min
Delta R.T.: 0.000 min
Response: 244160
Conc: 266.54 ng/ml



#29 AR-1254-4


R.T.: 7.015 min
Delta R.T.: 0.000 min
Response: 326697
Conc: 269.54 ng/ml



#30 AR-1254-5

R.T.: 8.419 min
Delta R.T.: 0.000 min
Response: 264502
Conc: 265.36 ng/ml

Instrument :
ECD_O
ClientSampleId :



#30 AR-1254-5

R.T.: 7.447 min
Delta R.T.: 0.000 min
Response: 441350
Conc: 266.74 ng/ml