

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011022\
 Data File : P0084134.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Jan 2022 00:10
 Operator : AJ\MA
 Sample : N1081-01
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 OB-CS-11

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/13/2022
 Supervised By :mohammad ahmed 01/13/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 02:54:17 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jan 10 11:26:21 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.046	4.037	2139257	456322	21.665	22.316
2) SA Decachlor...	11.248	9.442	1407221	375595	21.155	19.228
Target Compounds						
26) L6 AR-1254-1	7.292	6.208	36521	10231	10.419	8.577
27) L6 AR-1254-2	7.524	6.369	106118	10050	19.861	9.494 #
28) L6 AR-1254-3	7.910	6.796	254820	42588	45.257m	24.931m#
29) L6 AR-1254-4	8.207	7.038	171668	29231	41.895m	27.101m#
30) L6 AR-1254-5	8.643	7.474	198018	61222	46.152	41.008m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011022\
Data File : P0084134.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Jan 2022 00:10
Operator : AJ\MA
Sample : N1081-01
Misc :
ALS Vial : 42 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
OB-CS-11


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/13/2022
Supervised By :mohammad ahmed 01/13/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 02:54:17 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010722.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jan 10 11:26:21 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 5.046 min
Delta R.T.: 0.002 min
Response: 2139257
Conc: 21.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
OB-CS-11


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/13/2022
Supervised By :mohammad ahmed 01/13/2022




#1 Tetrachloro-m-xylene

R.T.: 4.037 min
Delta R.T.: 0.000 min
Response: 456322
Conc: 22.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.248 min
Delta R.T.: -0.010 min
Response: 1407221
Conc: 21.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.442 min
Delta R.T.: -0.007 min
Response: 375595
Conc: 19.23 ng/ml




#26 AR-1254-1

R.T.: 7.292 min
Delta R.T.: -0.001 min
Response: 36521
Conc: 10.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
OB-CS-11


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/13/2022
Supervised By :mohammad ahmed 01/13/2022




#26 AR-1254-1

R.T.: 6.208 min
Delta R.T.: -0.002 min
Response: 10231
Conc: 8.58 ng/ml




#27 AR-1254-2

R.T.: 7.524 min
Delta R.T.: -0.002 min
Response: 106118
Conc: 19.86 ng/ml



#27 AR-1254-2

R.T.: 6.369 min
Delta R.T.: -0.002 min
Response: 10050
Conc: 9.49 ng/ml




#28 AR-1254-3

R.T.: 7.910 min
Delta R.T.: 0.000 min
Response: 254820
Conc: 45.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
OB-CS-11


Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/13/2022
Supervised By :mohammad ahmed 01/13/2022




#28 AR-1254-3

R.T.: 6.796 min
Delta R.T.: -0.003 min
Response: 42588
Conc: 24.93 ng/ml m




#29 AR-1254-4

R.T.: 8.207 min
Delta R.T.: -0.001 min
Response: 171668
Conc: 41.89 ng/ml m



#29 AR-1254-4

R.T.: 7.038 min
Delta R.T.: -0.003 min
Response: 29231
Conc: 27.10 ng/ml m




#30 AR-1254-5

R.T.: 8.643 min
Delta R.T.: 0.000 min
Response: 198018
Conc: 46.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
OB-CS-11

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/13/2022
Supervised By :mohammad ahmed 01/13/2022



#30 AR-1254-5

R.T.: 7.474 min
Delta R.T.: -0.004 min
Response: 61222
Conc: 41.01 ng/ml m