

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011023\
 Data File : P0091928.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jan 2023 16:15
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 10 16:59:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 10 16:58:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.445	3.650	73682623	36397313	25.438	25.497
2)	SA Decachlor...	10.354	8.709	53416350	25391593	26.221	26.207
Target Compounds							
26)	L6 AR-1254-1	6.500	5.549	27762289	19569999	268.776	268.541
27)	L6 AR-1254-2	6.723	5.697	40199553	17479428	266.263	270.712
28)	L6 AR-1254-3	7.095	6.104	39533746	25541754	262.426	265.974
29)	L6 AR-1254-4	7.387	6.333	24907496	12840351	256.317	265.901
30)	L6 AR-1254-5	7.816	6.755	33040770	20556203	249.259	266.291



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011023\
Data File : PO091928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jan 2023 16:15
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 10 16:59:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011023.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 10 16:58:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.445 min
Delta R.T.: 0.000 min
Response: 73682623
Conc: 25.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.650 min
Delta R.T.: 0.000 min
Response: 36397313
Conc: 25.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.354 min
Delta R.T.: 0.000 min
Response: 53416350
Conc: 26.22 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.709 min
Delta R.T.: 0.000 min
Response: 25391593
Conc: 26.21 ng/ml



#26 AR-1254-1


R.T.: 6.500 min
Delta R.T.: 0.000 min
Response: 27762289
Conc: 268.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.549 min
Delta R.T.: 0.000 min
Response: 19569999
Conc: 268.54 ng/ml




#27 AR-1254-2

R.T.: 6.723 min
Delta R.T.: 0.000 min
Response: 40199553
Conc: 266.26 ng/ml



#27 AR-1254-2


R.T.: 5.697 min
Delta R.T.: 0.000 min
Response: 17479428
Conc: 270.71 ng/ml



#28 AR-1254-3


R.T.: 7.095 min
Delta R.T.: 0.000 min
Response: 39533746
Conc: 262.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.104 min
Delta R.T.: 0.000 min
Response: 25541754
Conc: 265.97 ng/ml




#29 AR-1254-4

R.T.: 7.387 min
Delta R.T.: 0.000 min
Response: 24907496
Conc: 256.32 ng/ml



#29 AR-1254-4


R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 12840351
Conc: 265.90 ng/ml



#30 AR-1254-5

R.T.: 7.816 min
Delta R.T.: 0.000 min
Response: 33040770
Conc: 249.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.755 min
Delta R.T.: 0.000 min
Response: 20556203
Conc: 266.29 ng/ml