

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011024\
 Data File : P0101053.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jan 2024 06:46
 Operator : YP/AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 07:37:59 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 07:28:00 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.342	3.638	18755786	12035192	4.722	4.673
2) SA Decachlor...	10.087	8.696	8953422	8179812	4.638	5.194
Target Compounds						
16) L4 AR-1242-1	5.511	4.733	4062488	3148180	50.519	55.541
17) L4 AR-1242-2	5.534	4.752	5875291	4245996	49.199	52.861
18) L4 AR-1242-3	5.596	4.931	3711518	2566023	47.263	60.327 #
19) L4 AR-1242-4	5.694	5.014	2874840	4489601	46.586	93.006 #
20) L4 AR-1242-5	6.434	5.543	5269415	5564284	91.465	106.158



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011024\
Data File : PO101053.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jan 2024 06:46
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 07:37:59 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 07:28:00 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.342 min
Delta R.T.: 0.000 min
Response: 18755786
Conc: 4.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 12035192
Conc: 4.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.087 min
Delta R.T.: -0.002 min
Response: 8953422
Conc: 4.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: -0.002 min
Response: 8179812
Conc: 5.19 ng/ml



#16 AR-1242-1


R.T.: 5.511 min
Delta R.T.: -0.001 min
Response: 4062488
Conc: 50.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 4.733 min
Delta R.T.: 0.000 min
Response: 3148180
Conc: 55.54 ng/ml




#17 AR-1242-2

R.T.: 5.534 min
Delta R.T.: -0.001 min
Response: 5875291
Conc: 49.20 ng/ml



#17 AR-1242-2


R.T.: 4.752 min
Delta R.T.: 0.000 min
Response: 4245996
Conc: 52.86 ng/ml



#18 AR-1242-3


R.T.: 5.596 min
Delta R.T.: 0.000 min
Response: 3711518
Conc: 47.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 4.931 min
Delta R.T.: -0.001 min
Response: 2566023
Conc: 60.33 ng/ml




#19 AR-1242-4

R.T.: 5.694 min
Delta R.T.: -0.001 min
Response: 2874840
Conc: 46.59 ng/ml



#19 AR-1242-4


R.T.: 5.014 min
Delta R.T.: 0.000 min
Response: 4489601
Conc: 93.01 ng/ml



#20 AR-1242-5

R.T.: 6.434 min
Delta R.T.: 0.000 min
Response: 5269415
Conc: 91.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 5.543 min
Delta R.T.: 0.000 min
Response: 5564284
Conc: 106.16 ng/ml