

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011024\
 Data File : P0101042.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jan 2024 03:41
 Operator : YP/AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 11 05:23:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 05:23:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.342	3.638	412.4E6	267.1E6	99.381	99.227
2) SA Decachlor...	10.086	8.696	199.8E6	161.6E6	98.877	98.266
Target Compounds						
3) L1 AR-1016-1	5.511	4.733	98845712	69751977	963.840	968.936
4) L1 AR-1016-2	5.534	4.753	150.6E6	100.6E6	983.255	976.681
5) L1 AR-1016-3	5.596	4.931	98715227	52514717	975.089	963.093
6) L1 AR-1016-4	5.695	4.972	77939682	46541584	982.169	953.019
7) L1 AR-1016-5	5.990	5.189	75552502	60855163	964.082	961.520
31) L7 AR-1260-1	7.122	6.231	137.3E6	112.1E6	986.203	973.235
32) L7 AR-1260-2	7.380	6.420	140.2E6	122.1E6	977.561	976.197
33) L7 AR-1260-3	7.741	6.575	110.7E6	117.9E6	988.854	991.297
34) L7 AR-1260-4	7.968	7.050	113.5E6	94054093	992.717	978.448
35) L7 AR-1260-5	8.285	7.293	208.7E6	197.1E6	996.787	990.271



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011024\
Data File : PO101042.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jan 2024 03:41
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 11 05:23:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 05:23:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.342 min
Delta R.T.: 0.000 min
Response: 412371961
Conc: 99.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.638 min
Delta R.T.: 0.000 min
Response: 267102352
Conc: 99.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.086 min
Delta R.T.: 0.000 min
Response: 199841686
Conc: 98.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.696 min
Delta R.T.: 0.000 min
Response: 161612371
Conc: 98.27 ng/ml



#3 AR-1016-1


R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 98845712
Conc: 963.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.733 min
Delta R.T.: 0.000 min
Response: 69751977
Conc: 968.94 ng/ml




#4 AR-1016-2

R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 150610976
Conc: 983.25 ng/ml



#4 AR-1016-2


R.T.: 4.753 min
Delta R.T.: 0.000 min
Response: 100648023
Conc: 976.68 ng/ml



#5 AR-1016-3


R.T.: 5.596 min
Delta R.T.: 0.000 min
Response: 98715227
Conc: 975.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 52514717
Conc: 963.09 ng/ml




#6 AR-1016-4

R.T.: 5.695 min
Delta R.T.: 0.000 min
Response: 77939682
Conc: 982.17 ng/ml



#6 AR-1016-4


R.T.: 4.972 min
Delta R.T.: 0.000 min
Response: 46541584
Conc: 953.02 ng/ml



#7 AR-1016-5


R.T.: 5.990 min
Delta R.T.: 0.000 min
Response: 75552502
Conc: 964.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.189 min
Delta R.T.: 0.000 min
Response: 60855163
Conc: 961.52 ng/ml




#31 AR-1260-1

R.T.: 7.122 min
Delta R.T.: 0.000 min
Response: 137279230
Conc: 986.20 ng/ml



#31 AR-1260-1


R.T.: 6.231 min
Delta R.T.: 0.000 min
Response: 112089896
Conc: 973.23 ng/ml



#32 AR-1260-2


R.T.: 7.380 min
Delta R.T.: 0.000 min
Response: 140169301
Conc: 977.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.420 min
Delta R.T.: 0.000 min
Response: 122110026
Conc: 976.20 ng/ml




#33 AR-1260-3

R.T.: 7.741 min
Delta R.T.: 0.000 min
Response: 110738438
Conc: 988.85 ng/ml



#33 AR-1260-3


R.T.: 6.575 min
Delta R.T.: 0.000 min
Response: 117918312
Conc: 991.30 ng/ml



#34 AR-1260-4


R.T.: 7.968 min
Delta R.T.: 0.000 min
Response: 113523522
Conc: 992.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.050 min
Delta R.T.: 0.000 min
Response: 94054093
Conc: 978.45 ng/ml



#35 AR-1260-5

R.T.: 8.285 min
Delta R.T.: 0.000 min
Response: 208653786
Conc: 996.79 ng/ml



#35 AR-1260-5

R.T.: 7.293 min
Delta R.T.: 0.000 min
Response: 197131666
Conc: 990.27 ng/ml