

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011121\
 Data File : P0074488.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Jan 2021 20:35
 Operator : AJ/MA
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC050

Manual Integrations
 APPROVED

Ankita
 1/12/2021 5:40:42 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 04:33:28 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 04:31:10 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.932	4.008	517230	292700	4.689	5.028
2)	SA Decachlor...	10.947	9.259	521871	296850	4.943	5.197
Target Compounds							
21)	L5 AR-1248-1	6.254	5.251	126352	76629	51.805	55.645
22)	L5 AR-1248-2	6.549	5.513	166294	107022	53.036	57.791
23)	L5 AR-1248-3	6.767	5.557	185585	104664	50.565	55.589
24)	L5 AR-1248-4	7.197	5.739	239001	125092	53.575	55.819
25)	L5 AR-1248-5	7.236	6.158	222970	127462	52.744	54.289m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011121\
Data File : PO074488.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Jan 2021 20:35
Operator : AJ/MA
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Ankita
1/12/2021 5:40:42 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 04:33:28 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 04:31:10 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.932 min
Delta R.T.: 0.000 min
Response: 517230
Conc: 4.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Ankita
1/12/2021 5:40:42 PM




#1 Tetrachloro-m-xylene

R.T.: 4.008 min
Delta R.T.: 0.000 min
Response: 292700
Conc: 5.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.947 min
Delta R.T.: 0.001 min
Response: 521871
Conc: 4.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.259 min
Delta R.T.: 0.000 min
Response: 296850
Conc: 5.20 ng/ml




#21 AR-1248-1

R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 126352
Conc: 51.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Ankita
1/12/2021 5:40:42 PM




#21 AR-1248-1

R.T.: 5.251 min
Delta R.T.: 0.000 min
Response: 76629
Conc: 55.65 ng/ml




#22 AR-1248-2

R.T.: 6.549 min
Delta R.T.: 0.000 min
Response: 166294
Conc: 53.04 ng/ml



#22 AR-1248-2

R.T.: 5.513 min
Delta R.T.: 0.000 min
Response: 107022
Conc: 57.79 ng/ml




#23 AR-1248-3

R.T.: 6.767 min
Delta R.T.: 0.000 min
Response: 185585
Conc: 50.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050


Manual Integrations
APPROVED

Ankita
1/12/2021 5:40:42 PM




#23 AR-1248-3

R.T.: 5.557 min
Delta R.T.: 0.000 min
Response: 104664
Conc: 55.59 ng/ml




#24 AR-1248-4

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 239001
Conc: 53.57 ng/ml



#24 AR-1248-4

R.T.: 5.739 min
Delta R.T.: 0.000 min
Response: 125092
Conc: 55.82 ng/ml




#25 AR-1248-5

R.T.: 7.236 min
Delta R.T.: 0.000 min
Response: 222970
Conc: 52.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Manual Integrations
APPROVED

Ankita
1/12/2021 5:40:42 PM



#25 AR-1248-5

R.T.: 6.158 min
Delta R.T.: 0.000 min
Response: 127462
Conc: 54.29 ng/ml m