

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011123\
 Data File : P0091950.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jan 2023 13:24
 Operator : YP/AJ
 Sample : 01081-06
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 31944

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/12/2023
 Supervised By :Ankita Jodhani 01/12/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 01:23:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 11 04:41:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.444	3.649	42181834	22590299	15.123	16.271
2) SA Decachlor...	10.351	8.703	27803984	11742216	14.202	13.243m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011123\
Data File : PO091950.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jan 2023 13:24
Operator : YP/AJ
Sample : 01081-06
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
31944


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/12/2023
Supervised By :Ankita Jodhani 01/12/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 01:23:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 11 04:41:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.444 min
Delta R.T.: 0.000 min
Response: 42181834
Conc: 15.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
31944


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/12/2023
Supervised By :Ankita Jodhani 01/12/2023




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: -0.001 min
Response: 22590299
Conc: 16.27 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.351 min
Delta R.T.: 0.001 min
Response: 27803984
Conc: 14.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.703 min
Delta R.T.: -0.008 min
Response: 11742216
Conc: 13.24 ng/ml m