

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011221\
 Data File : P0074527.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jan 2021 14:25
 Operator : AJ/MA
 Sample : M1084-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S1-06-B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 17:03:34 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 06:02:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.931	4.006	2302737	1274179	21.823	22.651
2) SA Decachlor...	10.946	9.258	1991522	1050450	19.622	19.132
Target Compounds						
26) L6 AR-1254-1	7.165	6.115	6043995	4526739	1374.216	1253.248
27) L6 AR-1254-2	7.395	6.274	13551135	6483988	1989.662	2080.510
28) L6 AR-1254-3	7.779	6.690	17421150	11876052	2483.652	2421.746
29) L6 AR-1254-4	8.075	6.928	17910328	9133489	2953.223	2633.814
30) L6 AR-1254-5	8.505	7.353	28732878	25883811	4946.323	5718.990



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011221\
Data File : PO074527.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Jan 2021 14:25
Operator : AJ/MA
Sample : M1084-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
S1-06-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 17:03:34 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 06:02:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 2302737
Conc: 21.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-06-B




#1 Tetrachloro-m-xylene

R.T.: 4.006 min
Delta R.T.: -0.004 min
Response: 1274179
Conc: 22.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.946 min
Delta R.T.: 0.000 min
Response: 1991522
Conc: 19.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.258 min
Delta R.T.: -0.003 min
Response: 1050450
Conc: 19.13 ng/ml



#26 AR-1254-1


R.T.: 7.165 min
Delta R.T.: -0.002 min
Response: 6043995
Conc: 1374.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-06-B




#26 AR-1254-1

R.T.: 6.115 min
Delta R.T.: -0.002 min
Response: 4526739
Conc: 1253.25 ng/ml




#27 AR-1254-2

R.T.: 7.395 min
Delta R.T.: -0.002 min
Response: 13551135
Conc: 1989.66 ng/ml



#27 AR-1254-2


R.T.: 6.274 min
Delta R.T.: -0.002 min
Response: 6483988
Conc: 2080.51 ng/ml



#28 AR-1254-3


R.T.: 7.779 min
Delta R.T.: -0.002 min
Response: 17421150
Conc: 2483.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-06-B




#28 AR-1254-3

R.T.: 6.690 min
Delta R.T.: -0.002 min
Response: 11876052
Conc: 2421.75 ng/ml




#29 AR-1254-4

R.T.: 8.075 min
Delta R.T.: -0.002 min
Response: 17910328
Conc: 2953.22 ng/ml



#29 AR-1254-4


R.T.: 6.928 min
Delta R.T.: -0.002 min
Response: 9133489
Conc: 2633.81 ng/ml



#30 AR-1254-5

R.T.: 8.505 min
Delta R.T.: -0.002 min
Response: 28732878
Conc: 4946.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
S1-06-B



#30 AR-1254-5

R.T.: 7.353 min
Delta R.T.: -0.002 min
Response: 25883811
Conc: 5718.99 ng/ml