

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011221\
 Data File : P0074540.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jan 2021 18:06
 Operator : AJ/MA
 Sample : M1085-01 5X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S2-01-B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 01:05:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011221.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 06:02:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.931	4.005	489596	279242	4.640	4.964
2)	SA Decachlor...	10.947	9.257	370721	241746	3.653	4.403
Target Compounds							
26)	L6 AR-1254-1	7.167	6.115	1196203	927768	271.979	256.857
27)	L6 AR-1254-2	7.396	6.274	2707719	1362123	397.564	437.063
28)	L6 AR-1254-3	7.780	6.690	3685052	2608128	525.361	531.845
29)	L6 AR-1254-4	8.076	6.928	3558517	1946745	586.762	561.381
30)	L6 AR-1254-5	8.506	7.354	6374987	6738077	1097.445	1488.768 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011221\
Data File : P0074540.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Jan 2021 18:06
Operator : AJ/MA
Sample : M1085-01 5X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
S2-01-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 01:05:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 06:02:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 489596
Conc: 4.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-01-B




#1 Tetrachloro-m-xylene

R.T.: 4.005 min
Delta R.T.: -0.005 min
Response: 279242
Conc: 4.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.947 min
Delta R.T.: 0.000 min
Response: 370721
Conc: 3.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.257 min
Delta R.T.: -0.004 min
Response: 241746
Conc: 4.40 ng/ml



#26 AR-1254-1


R.T.: 7.167 min
Delta R.T.: 0.000 min
Response: 1196203
Conc: 271.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-01-B




#26 AR-1254-1

R.T.: 6.115 min
Delta R.T.: -0.002 min
Response: 927768
Conc: 256.86 ng/ml




#27 AR-1254-2

R.T.: 7.396 min
Delta R.T.: -0.001 min
Response: 2707719
Conc: 397.56 ng/ml



#27 AR-1254-2


R.T.: 6.274 min
Delta R.T.: -0.002 min
Response: 1362123
Conc: 437.06 ng/ml



#28 AR-1254-3


R.T.: 7.780 min
Delta R.T.: -0.001 min
Response: 3685052
Conc: 525.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-01-B




#28 AR-1254-3

R.T.: 6.690 min
Delta R.T.: -0.002 min
Response: 2608128
Conc: 531.85 ng/ml




#29 AR-1254-4

R.T.: 8.076 min
Delta R.T.: 0.000 min
Response: 3558517
Conc: 586.76 ng/ml



#29 AR-1254-4


R.T.: 6.928 min
Delta R.T.: -0.002 min
Response: 1946745
Conc: 561.38 ng/ml



#30 AR-1254-5

R.T.: 8.506 min
Delta R.T.: 0.000 min
Response: 6374987
Conc: 1097.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-01-B



#30 AR-1254-5

R.T.: 7.354 min
Delta R.T.: -0.002 min
Response: 6738077
Conc: 1488.77 ng/ml