

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0011224\
 Data File : P0101156.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2024 20:23
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 23:13:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 11:52:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.341	3.637	78152426	49899497	19.294	18.929
2)	SA Decachlor...	10.081	8.692	45624932	37515405	23.101	22.716
Target Compounds							
3)	L1 AR-1016-1	5.510	4.730	132632	54315	1.276	0.730 #
4)	L1 AR-1016-2	5.534	4.741	27135	138278	0.176	1.324 #
5)	L1 AR-1016-3	5.593	4.938	162358	87561	1.607	1.591
6)	L1 AR-1016-4	5.693	4.976	95002	177242	1.208	3.492 #
7)	L1 AR-1016-5	5.996	5.181	37423	463591	0.473	7.159 #
8)	L2 AR-1221-1	4.555	3.863	49280	148275	1.033	4.788 #
9)	L2 AR-1221-2	4.649	3.939	1331872	811522	37.886	33.906
10)	L2 AR-1221-3	4.720	4.019	46170	45687	0.449	0.645 #
11)	L3 AR-1232-1	4.720	4.019	46170	45687	0.539	0.763 #
12)	L3 AR-1232-2	5.246	4.741	69108	138278	1.435	2.712 #
13)	L3 AR-1232-3	5.534	4.938	27135	87561	0.364	3.252 #
14)	L3 AR-1232-4	5.693	5.024	95002	105109	2.535	3.670 #
15)	L3 AR-1232-5	5.791	5.181	194050	463591	5.804	14.354 #
16)	L4 AR-1242-1	5.510	4.730	132632	54315	1.609	0.914 #
17)	L4 AR-1242-2	5.534	4.741	27135	138278	0.225	1.671 #
18)	L4 AR-1242-3	5.593	4.938	162358	87561	2.063	2.003
19)	L4 AR-1242-4	5.693	5.024	95002	105109	1.545	2.152 #
20)	L4 AR-1242-5	6.455	5.547	973327	145561	17.038	2.755 #
21)	L5 AR-1248-1	5.510	4.730	132632	54315	2.129	1.187 #
22)	L5 AR-1248-2	5.791	4.976	194050	177242	1.881	2.520 #
23)	L5 AR-1248-3	5.996	5.024	37423	105109	0.357	1.410 #
24)	L5 AR-1248-4	6.399	5.181	584903	463591	5.871	5.550
25)	L5 AR-1248-5	6.455	5.590	973327	178665	9.567	2.614 #
26)	L6 AR-1254-1	6.399	5.547	584903	145561	4.873	1.194 #
27)	L6 AR-1254-2	6.588	5.701	89840	274714	0.526	2.573 #
28)	L6 AR-1254-3	6.957	6.096	864816	167919	5.412	1.066 #
29)	L6 AR-1254-4	7.212	6.324	176252	166764	1.774	2.004
30)	L6 AR-1254-5	7.700	6.750	133121	399621	1.117	2.977 #
31)	L7 AR-1260-1	7.182f	6.226	1590954	247613	11.164	2.054 #
32)	L7 AR-1260-2	7.401	6.407	511630	474600	3.416	3.665
33)	L7 AR-1260-3	7.750	6.583	751589	20757	6.689	0.172 #
35)	L7 AR-1260-5	8.318	7.309	2560124	294890	11.984	1.497 #
36)	L8 AR-1262-1	7.700f	6.750	133121	399621	1.241	5.566 #
37)	L8 AR-1262-2	8.318	7.041	2560124	109127	10.344	0.825 #
38)	L8 AR-1262-3	8.593	7.535f	276264	7846	1.605	0.089 #
39)	L8 AR-1262-4	8.684	7.644	234856	301941	1.723	1.826
40)	L8 AR-1262-5	9.337	8.139	26708	469007	0.339	6.726 #
41)	L9 AR-1268-1	8.593	7.535f	276264	7846	0.895	0.030 #
42)	L9 AR-1268-2	8.684	7.644	234856	301941	0.852	1.262 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011224\
 Data File : P0101156.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2024 20:23
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 23:13:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 11 11:52:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	8.916	7.845	197271	659585	0.796	2.947 #
44) L9	AR-1268-4	9.337	8.139	26708	469007	0.322	6.071 #
45) L9	AR-1268-5	9.746	8.429	154091	474743	0.216	0.821 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011224\
Data File : PO101156.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2024 20:23
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 23:13:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 11 11:52:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.341 min
Delta R.T.: -0.002 min
Response: 78152426
Conc: 19.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.637 min
Delta R.T.: -0.001 min
Response: 49899497
Conc: 18.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.081 min
Delta R.T.: -0.007 min
Response: 45624932
Conc: 23.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.692 min
Delta R.T.: -0.005 min
Response: 37515405
Conc: 22.72 ng/ml



#3 AR-1016-1


R.T.: 5.510 min
Delta R.T.: -0.002 min
Response: 132632
Conc: 1.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#3 AR-1016-1

R.T.: 4.730 min
Delta R.T.: -0.004 min
Response: 54315
Conc: 0.73 ng/ml




#4 AR-1016-2

R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 27135
Conc: 0.18 ng/ml



#4 AR-1016-2


R.T.: 4.741 min
Delta R.T.: -0.012 min
Response: 138278
Conc: 1.32 ng/ml



#5 AR-1016-3


R.T.: 5.593 min
Delta R.T.: -0.004 min
Response: 162358
Conc: 1.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#5 AR-1016-3

R.T.: 4.938 min
Delta R.T.: 0.006 min
Response: 87561
Conc: 1.59 ng/ml




#6 AR-1016-4

R.T.: 5.693 min
Delta R.T.: -0.002 min
Response: 95002
Conc: 1.21 ng/ml



#6 AR-1016-4


R.T.: 4.976 min
Delta R.T.: 0.003 min
Response: 177242
Conc: 3.49 ng/ml



#7 AR-1016-5


R.T.: 5.996 min
Delta R.T.: 0.005 min
Response: 37423
Conc: 0.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#7 AR-1016-5

R.T.: 5.181 min
Delta R.T.: -0.008 min
Response: 463591
Conc: 7.16 ng/ml




#8 AR-1221-1

R.T.: 4.555 min
Delta R.T.: 0.004 min
Response: 49280
Conc: 1.03 ng/ml



#8 AR-1221-1


R.T.: 3.863 min
Delta R.T.: 0.010 min
Response: 148275
Conc: 4.79 ng/ml



#9 AR-1221-2


R.T.: 4.649 min
Delta R.T.: 0.011 min
Response: 1331872
Conc: 37.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 3.939 min
Delta R.T.: 0.000 min
Response: 811522
Conc: 33.91 ng/ml




#10 AR-1221-3

R.T.: 4.720 min
Delta R.T.: 0.006 min
Response: 46170
Conc: 0.45 ng/ml



#10 AR-1221-3


R.T.: 4.019 min
Delta R.T.: 0.003 min
Response: 45687
Conc: 0.65 ng/ml



#11 AR-1232-1


R.T.: 4.720 min
Delta R.T.: 0.005 min
Response: 46170
Conc: 0.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#11 AR-1232-1

R.T.: 4.019 min
Delta R.T.: 0.003 min
Response: 45687
Conc: 0.76 ng/ml




#12 AR-1232-2

R.T.: 5.246 min
Delta R.T.: 0.002 min
Response: 69108
Conc: 1.44 ng/ml



#12 AR-1232-2


R.T.: 4.741 min
Delta R.T.: -0.012 min
Response: 138278
Conc: 2.71 ng/ml



#13 AR-1232-3


R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 27135
Conc: 0.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#13 AR-1232-3

R.T.: 4.938 min
Delta R.T.: 0.006 min
Response: 87561
Conc: 3.25 ng/ml




#14 AR-1232-4

R.T.: 5.693 min
Delta R.T.: -0.002 min
Response: 95002
Conc: 2.53 ng/ml



#14 AR-1232-4


R.T.: 5.024 min
Delta R.T.: 0.009 min
Response: 105109
Conc: 3.67 ng/ml



#15 AR-1232-5


R.T.: 5.791 min
Delta R.T.: 0.004 min
Response: 194050
Conc: 5.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#15 AR-1232-5

R.T.: 5.181 min
Delta R.T.: -0.008 min
Response: 463591
Conc: 14.35 ng/ml




#16 AR-1242-1

R.T.: 5.510 min
Delta R.T.: 0.000 min
Response: 132632
Conc: 1.61 ng/ml



#16 AR-1242-1


R.T.: 4.730 min
Delta R.T.: -0.003 min
Response: 54315
Conc: 0.91 ng/ml



#17 AR-1242-2


R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 27135
Conc: 0.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#17 AR-1242-2

R.T.: 4.741 min
Delta R.T.: -0.011 min
Response: 138278
Conc: 1.67 ng/ml




#18 AR-1242-3

R.T.: 5.593 min
Delta R.T.: -0.002 min
Response: 162358
Conc: 2.06 ng/ml



#18 AR-1242-3


R.T.: 4.938 min
Delta R.T.: 0.007 min
Response: 87561
Conc: 2.00 ng/ml



#19 AR-1242-4


R.T.: 5.693 min
Delta R.T.: -0.001 min
Response: 95002
Conc: 1.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#19 AR-1242-4

R.T.: 5.024 min
Delta R.T.: 0.009 min
Response: 105109
Conc: 2.15 ng/ml




#20 AR-1242-5

R.T.: 6.455 min
Delta R.T.: 0.021 min
Response: 973327
Conc: 17.04 ng/ml



#20 AR-1242-5


R.T.: 5.547 min
Delta R.T.: 0.004 min
Response: 145561
Conc: 2.75 ng/ml



#21 AR-1248-1


R.T.: 5.510 min
Delta R.T.: 0.000 min
Response: 132632
Conc: 2.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#21 AR-1248-1

R.T.: 4.730 min
Delta R.T.: -0.003 min
Response: 54315
Conc: 1.19 ng/ml




#22 AR-1248-2

R.T.: 5.791 min
Delta R.T.: 0.005 min
Response: 194050
Conc: 1.88 ng/ml



#22 AR-1248-2


R.T.: 4.976 min
Delta R.T.: 0.004 min
Response: 177242
Conc: 2.52 ng/ml



#23 AR-1248-3


R.T.: 5.996 min
Delta R.T.: 0.007 min
Response: 37423
Conc: 0.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#23 AR-1248-3

R.T.: 5.024 min
Delta R.T.: 0.010 min
Response: 105109
Conc: 1.41 ng/ml




#24 AR-1248-4

R.T.: 6.399 min
Delta R.T.: 0.004 min
Response: 584903
Conc: 5.87 ng/ml



#24 AR-1248-4


R.T.: 5.181 min
Delta R.T.: -0.006 min
Response: 463591
Conc: 5.55 ng/ml



#25 AR-1248-5


R.T.: 6.455 min
Delta R.T.: 0.022 min
Response: 973327
Conc: 9.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#25 AR-1248-5

R.T.: 5.590 min
Delta R.T.: 0.006 min
Response: 178665
Conc: 2.61 ng/ml




#26 AR-1254-1

R.T.: 6.399 min
Delta R.T.: 0.028 min
Response: 584903
Conc: 4.87 ng/ml



#26 AR-1254-1


R.T.: 5.547 min
Delta R.T.: 0.002 min
Response: 145561
Conc: 1.19 ng/ml



#27 AR-1254-2


R.T.: 6.588 min
Delta R.T.: -0.002 min
Response: 89840
Conc: 0.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#27 AR-1254-2

R.T.: 5.701 min
Delta R.T.: 0.008 min
Response: 274714
Conc: 2.57 ng/ml




#28 AR-1254-3

R.T.: 6.957 min
Delta R.T.: -0.001 min
Response: 864816
Conc: 5.41 ng/ml



#28 AR-1254-3


R.T.: 6.096 min
Delta R.T.: -0.005 min
Response: 167919
Conc: 1.07 ng/ml



#29 AR-1254-4


R.T.: 7.212 min
Delta R.T.: -0.032 min
Response: 176252
Conc: 1.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#29 AR-1254-4

R.T.: 6.324 min
Delta R.T.: -0.007 min
Response: 166764
Conc: 2.00 ng/ml




#30 AR-1254-5

R.T.: 7.700 min
Delta R.T.: 0.034 min
Response: 133121
Conc: 1.12 ng/ml



#30 AR-1254-5


R.T.: 6.750 min
Delta R.T.: -0.002 min
Response: 399621
Conc: 2.98 ng/ml



#31 AR-1260-1


R.T.: 7.182 min
Delta R.T.: 0.060 min
Response: 1590954
Conc: 11.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#31 AR-1260-1

R.T.: 6.226 min
Delta R.T.: -0.005 min
Response: 247613
Conc: 2.05 ng/ml




#32 AR-1260-2

R.T.: 7.401 min
Delta R.T.: 0.021 min
Response: 511630
Conc: 3.42 ng/ml



#32 AR-1260-2


R.T.: 6.407 min
Delta R.T.: -0.013 min
Response: 474600
Conc: 3.66 ng/ml



#33 AR-1260-3


R.T.: 7.750 min
Delta R.T.: 0.008 min
Response: 751589
Conc: 6.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#33 AR-1260-3

R.T.: 6.583 min
Delta R.T.: 0.007 min
Response: 20757
Conc: 0.17 ng/ml




#34 AR-1260-4

R.T.: 7.926 min
Delta R.T.: -0.042 min
Response: -416728
Conc: N.D.



#34 AR-1260-4


R.T.: 7.041 min
Delta R.T.: -0.009 min
Response: 109127
Conc: 1.12 ng/ml



#35 AR-1260-5


R.T.: 8.318 min
Delta R.T.: 0.031 min
Response: 2560124
Conc: 11.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#35 AR-1260-5

R.T.: 7.309 min
Delta R.T.: 0.015 min
Response: 294890
Conc: 1.50 ng/ml




#36 AR-1262-1

R.T.: 7.700 min
Delta R.T.: 0.038 min
Response: 133121
Conc: 1.24 ng/ml



#36 AR-1262-1


R.T.: 6.750 min
Delta R.T.: 0.000 min
Response: 399621
Conc: 5.57 ng/ml



#37 AR-1262-2


R.T.: 8.318 min
Delta R.T.: 0.033 min
Response: 2560124
Conc: 10.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#37 AR-1262-2

R.T.: 7.041 min
Delta R.T.: -0.008 min
Response: 109127
Conc: 0.82 ng/ml




#38 AR-1262-3

R.T.: 8.593 min
Delta R.T.: -0.004 min
Response: 276264
Conc: 1.60 ng/ml



#38 AR-1262-3


R.T.: 7.535 min
Delta R.T.: -0.042 min
Response: 7846
Conc: 0.09 ng/ml



#39 AR-1262-4


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 234856
Conc: 1.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#39 AR-1262-4

R.T.: 7.644 min
Delta R.T.: 0.001 min
Response: 301941
Conc: 1.83 ng/ml




#40 AR-1262-5

R.T.: 9.337 min
Delta R.T.: -0.001 min
Response: 26708
Conc: 0.34 ng/ml



#40 AR-1262-5


R.T.: 8.139 min
Delta R.T.: -0.002 min
Response: 469007
Conc: 6.73 ng/ml



#41 AR-1268-1


R.T.: 8.593 min
Delta R.T.: 0.000 min
Response: 276264
Conc: 0.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#41 AR-1268-1

R.T.: 7.535 min
Delta R.T.: -0.043 min
Response: 7846
Conc: 0.03 ng/ml




#42 AR-1268-2

R.T.: 8.684 min
Delta R.T.: -0.002 min
Response: 234856
Conc: 0.85 ng/ml



#42 AR-1268-2


R.T.: 7.644 min
Delta R.T.: 0.001 min
Response: 301941
Conc: 1.26 ng/ml



#43 AR-1268-3


R.T.: 8.916 min
Delta R.T.: 0.002 min
Response: 197271
Conc: 0.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#43 AR-1268-3

R.T.: 7.845 min
Delta R.T.: -0.005 min
Response: 659585
Conc: 2.95 ng/ml




#44 AR-1268-4

R.T.: 9.337 min
Delta R.T.: 0.000 min
Response: 26708
Conc: 0.32 ng/ml



#44 AR-1268-4


R.T.: 8.139 min
Delta R.T.: -0.001 min
Response: 469007
Conc: 6.07 ng/ml



#45 AR-1268-5

R.T.: 9.746 min
Delta R.T.: -0.004 min
Response: 154091
Conc: 0.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK



#45 AR-1268-5

R.T.: 8.429 min
Delta R.T.: -0.007 min
Response: 474743
Conc: 0.82 ng/ml