

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0011322.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jan 14 00:41:00 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0084224.D 1000 =P0084223.D 750 =P0084223.D
 250 =P0084225.D 50 =P0084226.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.000	0.960	0.917	1.158	1.042	1.015	E5	9.07
2) SA	Decachlorobiphenyl	5.827	5.418	5.614	6.091	5.908	5.771	E4	4.53
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1	2.445	2.286	2.165	2.842	2.554	2.458	E3	10.61
17) L4	AR-1242-2	3.442	3.289	3.178	4.090	3.650	3.530	E3	10.19
18) L4	AR-1242-3	2.225	2.119	2.052	2.642	2.333	2.274	E3	10.17
19) L4	AR-1242-4	1.817	1.733	1.663	2.219	2.141	1.915	E3	13.05
20) L4	AR-1242-5	1.677	1.546	1.533	1.943	1.705	1.681	E3	9.84
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0084224.D 1000 =P0084222.D 750 =P0084223.D
 250 =P0084225.D 50 =P0084226.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.325	2.312	2.168	2.589	2.027	2.284	E4	9.16
2) SA	Decachlorobiphenyl	1.842	1.711	1.787	1.901	1.816	1.811	E4	3.88
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0011322.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jan 14 00:41:00 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0084224.D 1000 =P0084223.D 750 =P0084223.D
 250 =P0084225.D 50 =P0084226.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	6.181	5.979	5.788	7.066	5.889	6.180 E2	8.34
17) L4	AR-1242-2	8.763	8.260	8.046	9.603	8.062	8.547 E2	7.69
18) L4	AR-1242-3	4.545	4.396	4.250	5.182	4.597	4.594 E2	7.74
19) L4	AR-1242-4	4.456	4.268	4.143	5.129	4.510	4.501 E2	8.46
20) L4	AR-1242-5	5.437	5.120	5.087	6.096	5.343	5.417 E2	7.52
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range