

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0011322.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jan 14 01:59:42 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0084229.D 1000 =P0084227.D 750 =P0084228.D
 250 =P0084230.D 50 =P0084231.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.042	0.927	0.876	1.035	1.103	0.996	E5	9.29
2) SA	Decachlorobiphenyl	5.954	5.379	5.782	6.246	6.095	5.891	E4	5.67
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	1.909	1.663	1.634	1.924	1.962	1.818	E3	8.61
22) L5	AR-1248-2	2.806	2.355	2.335	2.855	2.840	2.638	E3	10.18
23) L5	AR-1248-3	2.932	2.494	2.475	2.946	3.236	2.817	E3	11.60
24) L5	AR-1248-4	3.020	2.600	2.589	3.090	3.481	2.956	E3	12.65
25) L5	AR-1248-5	2.842	2.495	2.483	2.903	3.180	2.780	E3	10.62
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0084229.D 1000 =P0084227.D 750 =P0084228.D
 250 =P0084230.D 50 =P0084231.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.465	2.261	2.120	2.307	2.231	2.277	E4	5.52
2) SA	Decachlorobiphenyl	1.855	1.714	1.813	1.914	1.833	1.826	E4	4.00
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0011322.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jan 14 01:59:42 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0084229.D 1000 =P0084227.D 750 =P0084228.D
 250 =P0084230.D 50 =P0084231.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	4.749	4.359	4.215	4.701	4.852	4.575 E2	5.98
22) L5	AR-1248-2	6.477	5.721	5.629	6.439	7.224	6.298 E2	10.32
23) L5	AR-1248-3	6.803	6.024	5.919	6.725	7.605	6.615 E2	10.31
24) L5	AR-1248-4	7.901	7.040	6.908	7.759	8.560	7.634 E2	8.84
25) L5	AR-1248-5	7.555	6.800	6.779	7.391	7.066	7.118 E2	4.89
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range