

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0011322.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jan 14 03:40:28 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0084234.D 1000 =P0084232.D 750 =P0084233.D
 250 =P0084235.D 50 =P0084236.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.025	0.962	0.927	0.967	1.230	1.022 E5	11.87
2) SA	Decachlorobiphenyl	6.030	5.494	5.874	6.335	6.411	6.029 E4	6.15
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	3.250	3.043	3.058	3.367	4.366	3.417 E3	16.03
27) L6	AR-1254-2	4.537	4.133	4.146	4.515	5.624	4.591 E3	13.26
28) L6	AR-1254-3	4.602	4.220	4.268	4.727	6.417	4.847 E3	18.65
29) L6	AR-1254-4	3.300	2.992	3.080	3.339	4.349	3.412 E3	15.94
30) L6	AR-1254-5	4.073	3.705	3.848	4.169	5.541	4.267 E3	17.23
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =P0084234.D 1000 =P0084232.D 750 =P0084233.D
 250 =P0084235.D 50 =P0084236.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.407	2.361	2.231	2.156	2.454	2.322 E4	5.36
2) SA	Decachlorobiphenyl	1.876	1.743	1.879	1.944	1.868	1.862 E4	3.92
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0011322.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jan 14 03:40:28 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0084234.D 1000 =P0084232.D 750 =P0084233.D
 250 =P0084235.D 50 =P0084236.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.157	1.076	1.086	1.118	1.348	1.157 E3	9.61
27) L6	AR-1254-2	0.990	0.914	0.934	0.970	1.156	0.993 E3	9.67
28) L6	AR-1254-3	1.587	1.490	1.516	1.565	1.862	1.604 E3	9.30
29) L6	AR-1254-4	1.002	0.932	0.962	1.013	1.137	1.009 E3	7.79
30) L6	AR-1254-5	1.342	1.251	1.309	1.347	1.625	1.375 E3	10.54
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range