

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011420\
 Data File : P0065483.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jan 2020 13:25
 Operator : DD\AJ
 Sample : L1004-07
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BU-RO-229

Manual Integrations
 APPROVED

mohammad
 1/15/2020 4:15:10 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 13:51:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.171	3.445	700179	835755	38.344m	37.993
2) SA Decachlor...	9.677	8.349	646538	685582	19.082	16.532

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011420\
Data File : P0065483.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Jan 2020 13:25
Operator : DD\AJ
Sample : L1004-07
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BU-RO-229


Manual Integrations
APPROVED

mohammad
1/15/2020 4:15:10 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 13:51:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.171 min
Delta R.T.: 0.004 min
Response: 700179
Conc: 38.34 ng/ml m

Instrument :
ECD_O
ClientSampleId :
BU-RO-229


Manual Integrations
APPROVED

mohammad
1/15/2020 4:15:10 PM




#1 Tetrachloro-m-xylene

R.T.: 3.445 min
Delta R.T.: 0.003 min
Response: 835755
Conc: 37.99 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.677 min
Delta R.T.: 0.000 min
Response: 646538
Conc: 19.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.349 min
Delta R.T.: -0.003 min
Response: 685582
Conc: 16.53 ng/ml