

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011520\
 Data File : P0065513.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Jan 2020 15:44
 Operator : DD\AJ
 Sample : L1123-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-01-COMP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 15 15:58:38 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.165	3.440	518944	597949	28.419	27.183
2) SA Decachlor...	9.669	8.343	615706	709317	18.172	17.104
Target Compounds						
3) L1 AR-1016-1	5.251	0.000	12427	0	14.249	N.D. #
16) L4 AR-1242-1	5.251	0.000	12427	0	21.346	N.D. #
21) L5 AR-1248-1	5.251	0.000	12427	0	26.708	N.D. #
27) L6 AR-1254-2	6.375	5.434	13056	6441	9.501	4.869 #
28) L6 AR-1254-3	6.737	5.831	17925	34638	11.565	15.370 #
29) L6 AR-1254-4	7.022	6.059	40271	16093	31.592	10.674 #
30) L6 AR-1254-5	7.436	6.472	20224	35930	14.401	16.141
31) L7 AR-1260-1	6.900	5.959	24499	14641	17.763	7.777 #
32) L7 AR-1260-2	7.226	6.147	17941	14494	10.225	5.966 #
33) L7 AR-1260-3	0.000	6.297	0	17419	N.D.	8.047 #
34) L7 AR-1260-4	7.723	6.802	19653	14793	11.256	7.374 #
45) L9 AR-1268-5	9.371	0.000	17095	0	1.543	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011520\
Data File : P0065513.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Jan 2020 15:44
Operator : DD\AJ
Sample : L1123-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-01-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 15 15:58:38 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.165 min
Delta R.T.: -0.002 min
Response: 518944
Conc: 28.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-01-COMP




#1 Tetrachloro-m-xylene

R.T.: 3.440 min
Delta R.T.: -0.002 min
Response: 597949
Conc: 27.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.669 min
Delta R.T.: -0.009 min
Response: 615706
Conc: 18.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.343 min
Delta R.T.: -0.009 min
Response: 709317
Conc: 17.10 ng/ml



#3 AR-1016-1


R.T.: 5.251 min
Delta R.T.: -0.068 min
Response: 12427
Conc: 14.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-01-COMP




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.503 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.251 min
Delta R.T.: -0.067 min
Response: 12427
Conc: 21.35 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.502 min
Response: 0
Conc: N.D.



#21 AR-1248-1


R.T.: 5.251 min
Delta R.T.: -0.069 min
Response: 12427
Conc: 26.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-01-COMP




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.503 min
Response: 0
Conc: N.D.




#27 AR-1254-2

R.T.: 6.375 min
Delta R.T.: -0.003 min
Response: 13056
Conc: 9.50 ng/ml



#27 AR-1254-2


R.T.: 5.434 min
Delta R.T.: -0.005 min
Response: 6441
Conc: 4.87 ng/ml



#28 AR-1254-3


R.T.: 6.737 min
Delta R.T.: -0.004 min
Response: 17925
Conc: 11.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-01-COMP




#28 AR-1254-3

R.T.: 5.831 min
Delta R.T.: -0.006 min
Response: 34638
Conc: 15.37 ng/ml




#29 AR-1254-4

R.T.: 7.022 min
Delta R.T.: -0.004 min
Response: 40271
Conc: 31.59 ng/ml



#29 AR-1254-4


R.T.: 6.059 min
Delta R.T.: -0.005 min
Response: 16093
Conc: 10.67 ng/ml



#30 AR-1254-5


R.T.: 7.436 min
Delta R.T.: -0.004 min
Response: 20224
Conc: 14.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-01-COMP




#30 AR-1254-5

R.T.: 6.472 min
Delta R.T.: -0.006 min
Response: 35930
Conc: 16.14 ng/ml




#31 AR-1260-1

R.T.: 6.900 min
Delta R.T.: -0.003 min
Response: 24499
Conc: 17.76 ng/ml



#31 AR-1260-1


R.T.: 5.959 min
Delta R.T.: -0.006 min
Response: 14641
Conc: 7.78 ng/ml



#32 AR-1260-2


R.T.: 7.226 min
Delta R.T.: 0.067 min
Response: 17941
Conc: 10.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-01-COMP




#32 AR-1260-2

R.T.: 6.147 min
Delta R.T.: -0.006 min
Response: 14494
Conc: 5.97 ng/ml




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 7.514 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 6.297 min
Delta R.T.: -0.007 min
Response: 17419
Conc: 8.05 ng/ml



#34 AR-1260-4


R.T.: 7.723 min
Delta R.T.: -0.015 min
Response: 19653
Conc: 11.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-01-COMP




#34 AR-1260-4

R.T.: 6.802 min
Delta R.T.: 0.031 min
Response: 14793
Conc: 7.37 ng/ml



#45 AR-1268-5

R.T.: 9.371 min
Delta R.T.: -0.008 min
Response: 17095
Conc: 1.54 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 8.120 min
Response: 0
Conc: N.D.