

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011620\
 Data File : P0065558.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Jan 2020 16:25
 Operator : DD\AJ
 Sample : L1125-06
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 FEQCFB-01-14-20

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 17:04:06 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.159	3.434	411065	452458	22.511	20.569
2)	SA Decachlor...	9.661	8.337	631151	776934	18.628	18.735
Target Compounds							
36)	L8 AR-1262-1	0.000	6.572	0	14178	N.D.	15.039 #
45)	L9 AR-1268-5	0.000	8.154	0	51200	N.D.	3.561 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011620\
Data File : P0065558.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Jan 2020 16:25
Operator : DD\AJ
Sample : L1125-06
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
FEQCFB-01-14-20

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 17:04:06 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.159 min
Delta R.T.: -0.008 min
Response: 411065
Conc: 22.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
FEQCFB-01-14-20




#1 Tetrachloro-m-xylene

R.T.: 3.434 min
Delta R.T.: -0.007 min
Response: 452458
Conc: 20.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.661 min
Delta R.T.: -0.017 min
Response: 631151
Conc: 18.63 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.337 min
Delta R.T.: -0.016 min
Response: 776934
Conc: 18.73 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T.: 7.515 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
FEQCFB-01-14-20




#36 AR-1262-1

R.T.: 6.572 min
Delta R.T.: 0.004 min
Response: 14178
Conc: 15.04 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.380 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.154 min
Delta R.T.: 0.034 min
Response: 51200
Conc: 3.56 ng/ml