

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011623\
 Data File : P0092039.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jan 2023 13:53
 Operator : YP/AJ
 Sample : 01155-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 34254

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/17/2023
 Supervised By :Ankita Jodhani 01/17/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 20:44:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 11 04:41:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.444	3.647	52921564	26208756	18.974m	18.877
2) SA Decachlor...	10.362	8.703	45855191	21298433	23.423	24.020

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011623\
Data File : PO092039.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jan 2023 13:53
Operator : YP/AJ
Sample : 01155-01
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
34254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/17/2023
Supervised By :Ankita Jodhani 01/17/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 20:44:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 11 04:41:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.444 min
Delta R.T.: 0.000 min
Response: 52921564
Conc: 18.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
34254


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/17/2023
Supervised By :Ankita Jodhani 01/17/2023




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: -0.004 min
Response: 26208756
Conc: 18.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.362 min
Delta R.T.: 0.012 min
Response: 45855191
Conc: 23.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.703 min
Delta R.T.: -0.007 min
Response: 21298433
Conc: 24.02 ng/ml