

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011720\
 Data File : P0065627.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jan 2020 19:14
 Operator : DD\AJ
 Sample : L1162-10
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MH-720

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 23:20:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.170	3.444	346715	440229	18.987	20.013
2) SA Decachlor...	9.669	8.342	713855	821238	21.069	19.803

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011720\
Data File : P0065627.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jan 2020 19:14
Operator : DD\AJ
Sample : L1162-10
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MH-720

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 23:20:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.170 min
Delta R.T.: 0.003 min
Response: 346715
Conc: 18.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
MH-720




#1 Tetrachloro-m-xylene

R.T.: 3.444 min
Delta R.T.: 0.003 min
Response: 440229
Conc: 20.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.669 min
Delta R.T.: -0.009 min
Response: 713855
Conc: 21.07 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.342 min
Delta R.T.: -0.010 min
Response: 821238
Conc: 19.80 ng/ml