

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011720\
 Data File : P0065649.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2020 1:55
 Operator : DD\AJ
 Sample : L1008-08
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CL-01-011620

Manual Integrations
 APPROVED

mohammad
 1/20/2020 10:13:49 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 04:38:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.167	3.441	562206	716981	30.788m	32.594m
2) SA Decachlor...	9.668	8.342	695753	883931	20.534	21.315

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011720\
Data File : P0065649.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jan 2020 1:55
Operator : DD\AJ
Sample : L1008-08
Misc :
ALS Vial : 45 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CL-01-011620


Manual Integrations
APPROVED

mohammad
1/20/2020 10:13:49 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 04:38:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.167 min
Delta R.T.: 0.000 min
Response: 562206
Conc: 30.79 ng/ml m

Instrument :
ECD_O
ClientSampleId :
CL-01-011620


Manual Integrations
APPROVED

mohammad
1/20/2020 10:13:49 AM




#1 Tetrachloro-m-xylene

R.T.: 3.441 min
Delta R.T.: 0.000 min
Response: 716981
Conc: 32.59 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.668 min
Delta R.T.: -0.010 min
Response: 695753
Conc: 20.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.342 min
Delta R.T.: -0.011 min
Response: 883931
Conc: 21.31 ng/ml