

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0011722.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 18 02:51:20 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0084295.D 1000 =P0084293.D 750 =P0084294.D
 250 =P0084296.D 50 =P0084297.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	7.847	7.591	7.561	7.925	9.473	8.080 E4	9.84
2) SA	Decachlorobiphenyl	5.024	4.495	4.772	5.151	5.686	5.026 E4	8.88
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	2.928	2.479	2.690	2.853	3.630	2.916 E3	14.90
27) L6	AR-1254-2	3.847	3.510	3.647	3.958	4.710	3.934 E3	11.87
28) L6	AR-1254-3	3.953	3.583	3.717	4.147	5.323	4.145 E3	16.73
29) L6	AR-1254-4	2.704	2.426	2.556	2.857	3.284	2.765 E3	12.00
30) L6	AR-1254-5	3.312	3.026	3.172	3.405	3.653	3.314 E3	7.18
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =P0084295.D 1000 =P0084293.D 750 =P0084294.D
 250 =P0084296.D 50 =P0084297.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.424	2.434	2.398	2.326	2.576	2.432 E4	3.75
2) SA	Decachlorobiphenyl	1.925	1.753	1.837	1.987	2.018	1.904 E4	5.75
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0011722.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 18 02:51:20 2022
 Response Via : Initial Calibration

Calibration Files

500 =P0084295.D 1000 =P0084293.D 750 =P0084294.D
 250 =P0084296.D 50 =P0084297.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.229	1.151	1.175	1.240	1.445	1.248 E3	9.32
27) L6	AR-1254-2	1.060	0.971	1.009	1.068	1.266	1.075 E3	10.60
28) L6	AR-1254-3	1.690	1.587	1.628	1.722	2.040	1.733 E3	10.36
29) L6	AR-1254-4	1.017	0.932	0.973	1.185	1.171	1.056 E3	10.98
30) L6	AR-1254-5	1.428	1.318	1.364	1.454	1.737	1.460 E3	11.22
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range