

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011722\
 Data File : P0084288.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jan 2022 21:20
 Operator : AJ\MA
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 01:58:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 18 01:55:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	6.521	5.306	7003008	2270151	94.904	97.490
2) SA Decachlor...	13.783	11.320	4532545	1767953	96.641	96.653
Target Compounds						
21) L5 AR-1248-1	7.973	6.776	1317647	443478	951.670	973.894
22) L5 AR-1248-2	8.288	7.069	1895107	598425	939.416	957.852
23) L5 AR-1248-3	8.523	7.120	2072193	630978	946.211	961.199
24) L5 AR-1248-4	8.971	7.326	2142066	737257	948.131	961.125
25) L5 AR-1248-5	9.015	7.785	2057184	709127	954.439	966.133



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011722\
Data File : P0084288.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jan 2022 21:20
Operator : AJ\MA
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 01:58:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 18 01:55:06 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 6.521 min
Delta R.T.: 0.000 min
Response: 7003008
Conc: 94.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 5.306 min
Delta R.T.: 0.000 min
Response: 2270151
Conc: 97.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 13.783 min
Delta R.T.: 0.000 min
Response: 4532545
Conc: 96.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 11.320 min
Delta R.T.: 0.000 min
Response: 1767953
Conc: 96.65 ng/ml



#21 AR-1248-1


R.T.: 7.973 min
Delta R.T.: 0.000 min
Response: 1317647
Conc: 951.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 6.776 min
Delta R.T.: 0.000 min
Response: 443478
Conc: 973.89 ng/ml




#22 AR-1248-2

R.T.: 8.288 min
Delta R.T.: 0.000 min
Response: 1895107
Conc: 939.42 ng/ml



#22 AR-1248-2


R.T.: 7.069 min
Delta R.T.: 0.000 min
Response: 598425
Conc: 957.85 ng/ml



#23 AR-1248-3


R.T.: 8.523 min
Delta R.T.: 0.000 min
Response: 2072193
Conc: 946.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 7.120 min
Delta R.T.: 0.000 min
Response: 630978
Conc: 961.20 ng/ml




#24 AR-1248-4

R.T.: 8.971 min
Delta R.T.: 0.000 min
Response: 2142066
Conc: 948.13 ng/ml



#24 AR-1248-4


R.T.: 7.326 min
Delta R.T.: 0.000 min
Response: 737257
Conc: 961.12 ng/ml



#25 AR-1248-5

R.T.: 9.015 min
Delta R.T.: 0.000 min
Response: 2057184
Conc: 954.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 7.785 min
Delta R.T.: 0.000 min
Response: 709127
Conc: 966.13 ng/ml