

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011723\
 Data File : P0092098.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jan 2023 13:42
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 17 20:31:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 11 04:41:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.447	3.647	118.6E6	57251074	42.512	41.236
2) SA Decachlor...	10.368	8.702	95711527	48401960	48.890	54.587
Target Compounds						
26) L6 AR-1254-1	6.504	5.544	43801931	31891039	423.364	431.957
27) L6 AR-1254-2	6.729	5.692	63533086	28396769	419.137	434.567
28) L6 AR-1254-3	7.101	6.099	65210445	43529716	436.088	448.475
29) L6 AR-1254-4	7.394	6.329	44123991	23523692	464.255	485.371
30) L6 AR-1254-5	7.822	6.750	58733146	37215800	473.453	483.297


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011723\
Data File : PO092098.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jan 2023 13:42
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 17 20:31:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 11 04:41:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: 0.002 min
Response: 118573744
Conc: 42.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: -0.003 min
Response: 57251074
Conc: 41.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.368 min
Delta R.T.: 0.018 min
Response: 95711527
Conc: 48.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.702 min
Delta R.T.: -0.008 min
Response: 48401960
Conc: 54.59 ng/ml



#26 AR-1254-1


R.T.: 6.504 min
Delta R.T.: 0.005 min
Response: 43801931
Conc: 423.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.544 min
Delta R.T.: -0.004 min
Response: 31891039
Conc: 431.96 ng/ml




#27 AR-1254-2

R.T.: 6.729 min
Delta R.T.: 0.006 min
Response: 63533086
Conc: 419.14 ng/ml



#27 AR-1254-2


R.T.: 5.692 min
Delta R.T.: -0.005 min
Response: 28396769
Conc: 434.57 ng/ml



#28 AR-1254-3


R.T.: 7.101 min
Delta R.T.: 0.007 min
Response: 65210445
Conc: 436.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.099 min
Delta R.T.: -0.004 min
Response: 43529716
Conc: 448.47 ng/ml




#29 AR-1254-4

R.T.: 7.394 min
Delta R.T.: 0.007 min
Response: 44123991
Conc: 464.26 ng/ml



#29 AR-1254-4


R.T.: 6.329 min
Delta R.T.: -0.004 min
Response: 23523692
Conc: 485.37 ng/ml



#30 AR-1254-5

R.T.: 7.822 min
Delta R.T.: 0.008 min
Response: 58733146
Conc: 473.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.750 min
Delta R.T.: -0.004 min
Response: 37215800
Conc: 483.30 ng/ml