

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011821\
 Data File : P0074863.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2021 12:27
 Operator : AJ/MA
 Sample : M1141-22MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ETGI-354MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 14:34:27 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 06:02:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.930	4.004	2143329	1101440	20.312	19.580
2)	SA Decachlor...	10.951	9.258	2026409	1020906	19.966	18.594
Target Compounds							
3)	L1 AR-1016-1	6.254	5.248	1723950	927255	421.655	405.397
4)	L1 AR-1016-2	6.278	5.268	2438722	1288683	428.226	407.840
5)	L1 AR-1016-3	6.344	5.458	1475435	704054	434.782	410.157
6)	L1 AR-1016-4	6.452	5.510	1259914	574571	443.120	408.826
7)	L1 AR-1016-5	6.768	5.737	1144205	707590	418.352	411.487
31)	L7 AR-1260-1	7.948	6.825	2296563	1366003	460.350	428.176
32)	L7 AR-1260-2	8.214	7.022	3380927	1685150	497.041	396.408
33)	L7 AR-1260-3	8.581	7.174	2200132	1496323	383.851	415.245
34)	L7 AR-1260-4	8.811	7.653	2273800	1128916	415.586	360.524
35)	L7 AR-1260-5	9.138	7.900	4990781	2739059	399.423	344.594



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011821\
Data File : P0074863.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jan 2021 12:27
Operator : AJ/MA
Sample : M1141-22MSD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleID :
ETGI-354MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 14:34:27 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 06:02:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.930 min
Delta R.T.: 0.000 min
Response: 2143329
Conc: 20.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
ETGI-354MSD




#1 Tetrachloro-m-xylene

R.T.: 4.004 min
Delta R.T.: -0.006 min
Response: 1101440
Conc: 19.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.951 min
Delta R.T.: 0.005 min
Response: 2026409
Conc: 19.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.258 min
Delta R.T.: -0.003 min
Response: 1020906
Conc: 18.59 ng/ml



#3 AR-1016-1


R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 1723950
Conc: 421.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
ETGI-354MSD




#3 AR-1016-1

R.T.: 5.248 min
Delta R.T.: -0.005 min
Response: 927255
Conc: 405.40 ng/ml




#4 AR-1016-2

R.T.: 6.278 min
Delta R.T.: 0.000 min
Response: 2438722
Conc: 428.23 ng/ml



#4 AR-1016-2


R.T.: 5.268 min
Delta R.T.: -0.005 min
Response: 1288683
Conc: 407.84 ng/ml



#5 AR-1016-3


R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 1475435
Conc: 434.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
ETGI-354MSD




#5 AR-1016-3

R.T.: 5.458 min
Delta R.T.: -0.005 min
Response: 704054
Conc: 410.16 ng/ml




#6 AR-1016-4

R.T.: 6.452 min
Delta R.T.: 0.000 min
Response: 1259914
Conc: 443.12 ng/ml



#6 AR-1016-4


R.T.: 5.510 min
Delta R.T.: -0.005 min
Response: 574571
Conc: 408.83 ng/ml



#7 AR-1016-5


R.T.: 6.768 min
Delta R.T.: 0.000 min
Response: 1144205
Conc: 418.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
ETGI-354MSD




#7 AR-1016-5

R.T.: 5.737 min
Delta R.T.: -0.005 min
Response: 707590
Conc: 411.49 ng/ml




#31 AR-1260-1

R.T.: 7.948 min
Delta R.T.: 0.002 min
Response: 2296563
Conc: 460.35 ng/ml



#31 AR-1260-1


R.T.: 6.825 min
Delta R.T.: -0.004 min
Response: 1366003
Conc: 428.18 ng/ml



#32 AR-1260-2


R.T.: 8.214 min
Delta R.T.: 0.002 min
Response: 3380927
Conc: 497.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
ETGI-354MSD




#32 AR-1260-2

R.T.: 7.022 min
Delta R.T.: -0.003 min
Response: 1685150
Conc: 396.41 ng/ml




#33 AR-1260-3

R.T.: 8.581 min
Delta R.T.: 0.003 min
Response: 2200132
Conc: 383.85 ng/ml



#33 AR-1260-3


R.T.: 7.174 min
Delta R.T.: -0.005 min
Response: 1496323
Conc: 415.24 ng/ml



#34 AR-1260-4


R.T.: 8.811 min
Delta R.T.: 0.002 min
Response: 2273800
Conc: 415.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
ETGI-354MSD




#34 AR-1260-4

R.T.: 7.653 min
Delta R.T.: -0.004 min
Response: 1128916
Conc: 360.52 ng/ml



#35 AR-1260-5

R.T.: 9.138 min
Delta R.T.: 0.003 min
Response: 4990781
Conc: 399.42 ng/ml



#35 AR-1260-5

R.T.: 7.900 min
Delta R.T.: -0.004 min
Response: 2739059
Conc: 344.59 ng/ml