

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011821\
 Data File : P0074891.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jan 2021 20:27
 Operator : AJ/MA
 Sample : M1085-03DL 2X
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S2-01-CDL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 19 01:06:43 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 06:02:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.928	4.002	1181562	640560	11.198	11.387
2) SA Decachlor...	10.946	9.254	976954	558097	9.626	10.165
Target Compounds						
26) L6 AR-1254-1	7.165	6.111	1532317	1015761	348.401	281.218
27) L6 AR-1254-2	7.395	6.270	2865626	1277304	420.749	409.847
28) L6 AR-1254-3	7.779	6.686	3647159	2407552	519.958	490.944
29) L6 AR-1254-4	8.075	6.924	3267670	1644265	538.804	474.155
30) L6 AR-1254-5	8.504	7.350	5009394	6059878	862.360	1338.921 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011821\
Data File : PO074891.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jan 2021 20:27
Operator : AJ/MA
Sample : M1085-03DL 2X
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
S2-01-CDL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 19 01:06:43 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 06:02:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.928 min
Delta R.T.: -0.003 min
Response: 1181562
Conc: 11.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-01-CDL




#1 Tetrachloro-m-xylene

R.T.: 4.002 min
Delta R.T.: -0.008 min
Response: 640560
Conc: 11.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.946 min
Delta R.T.: 0.000 min
Response: 976954
Conc: 9.63 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.254 min
Delta R.T.: -0.007 min
Response: 558097
Conc: 10.16 ng/ml



#26 AR-1254-1


R.T.: 7.165 min
Delta R.T.: -0.002 min
Response: 1532317
Conc: 348.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-01-CDL




#26 AR-1254-1

R.T.: 6.111 min
Delta R.T.: -0.006 min
Response: 1015761
Conc: 281.22 ng/ml




#27 AR-1254-2

R.T.: 7.395 min
Delta R.T.: -0.003 min
Response: 2865626
Conc: 420.75 ng/ml



#27 AR-1254-2


R.T.: 6.270 min
Delta R.T.: -0.005 min
Response: 1277304
Conc: 409.85 ng/ml



#28 AR-1254-3


R.T.: 7.779 min
Delta R.T.: -0.002 min
Response: 3647159
Conc: 519.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-01-CDL




#28 AR-1254-3

R.T.: 6.686 min
Delta R.T.: -0.006 min
Response: 2407552
Conc: 490.94 ng/ml




#29 AR-1254-4

R.T.: 8.075 min
Delta R.T.: -0.002 min
Response: 3267670
Conc: 538.80 ng/ml



#29 AR-1254-4


R.T.: 6.924 min
Delta R.T.: -0.006 min
Response: 1644265
Conc: 474.15 ng/ml



#30 AR-1254-5

R.T.: 8.504 min
Delta R.T.: -0.002 min
Response: 5009394
Conc: 862.36 ng/ml

Instrument :
ECD_O
ClientSampleId :
S2-01-CDL



#30 AR-1254-5

R.T.: 7.350 min
Delta R.T.: -0.006 min
Response: 6059878
Conc: 1338.92 ng/ml