

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011823\
 Data File : P0092118.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2023 12:20
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 18 16:30:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 11 04:41:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.448	3.647	134.5E6	62295165	48.222	44.869
2) SA Decachlor...	10.368	8.699	101.8E6	48256584	51.979	54.423
Target Compounds						
26) L6 AR-1254-1	6.505	5.543	50292383	34943013	486.097	473.295
27) L6 AR-1254-2	6.730	5.692	73028506	30941803	481.780	473.515
28) L6 AR-1254-3	7.102	6.097	72463472	47682867	484.592	491.263
29) L6 AR-1254-4	7.394	6.328	47643541	26790728	501.287	552.780
30) L6 AR-1254-5	7.823	6.748	57526481	39587599	463.726	514.098



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011823\
Data File : PO092118.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2023 12:20
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 18 16:30:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 11 04:41:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.448 min
Delta R.T.: 0.003 min
Response: 134501320
Conc: 48.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: -0.003 min
Response: 62295165
Conc: 44.87 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.368 min
Delta R.T.: 0.018 min
Response: 101757938
Conc: 51.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.699 min
Delta R.T.: -0.012 min
Response: 48256584
Conc: 54.42 ng/ml



#26 AR-1254-1


R.T.: 6.505 min
Delta R.T.: 0.006 min
Response: 50292383
Conc: 486.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.543 min
Delta R.T.: -0.005 min
Response: 34943013
Conc: 473.30 ng/ml




#27 AR-1254-2

R.T.: 6.730 min
Delta R.T.: 0.007 min
Response: 73028506
Conc: 481.78 ng/ml



#27 AR-1254-2


R.T.: 5.692 min
Delta R.T.: -0.006 min
Response: 30941803
Conc: 473.51 ng/ml



#28 AR-1254-3


R.T.: 7.102 min
Delta R.T.: 0.007 min
Response: 72463472
Conc: 484.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.097 min
Delta R.T.: -0.006 min
Response: 47682867
Conc: 491.26 ng/ml




#29 AR-1254-4

R.T.: 7.394 min
Delta R.T.: 0.008 min
Response: 47643541
Conc: 501.29 ng/ml



#29 AR-1254-4


R.T.: 6.328 min
Delta R.T.: -0.005 min
Response: 26790728
Conc: 552.78 ng/ml



#30 AR-1254-5

R.T.: 7.823 min
Delta R.T.: 0.009 min
Response: 57526481
Conc: 463.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.748 min
Delta R.T.: -0.006 min
Response: 39587599
Conc: 514.10 ng/ml