

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011823\
 Data File : P0092132.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jan 2023 17:13
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 19 03:17:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 11 04:41:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.449	3.649	127.0E6	59137542	45.524	42.595
2)	SA Decachlor...	10.365	8.697	97808769	46038815	49.961	51.922
Target Compounds							
21)	L5 AR-1248-1	5.634	4.738	23917981	12819433	437.015	449.515
22)	L5 AR-1248-2	5.913	4.976	39266997	19375926	453.963	436.842
23)	L5 AR-1248-3	6.120	5.018	42886801	20180204	466.153	442.953
24)	L5 AR-1248-4	6.532	5.189	44207981	22823471	497.006	437.718
25)	L5 AR-1248-5	6.571	5.584	42740585	20562696	494.054	464.924



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011823\
Data File : PO092132.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jan 2023 17:13
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 19 03:17:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 11 04:41:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: 0.004 min
Response: 126974878
Conc: 45.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: -0.002 min
Response: 59137542
Conc: 42.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.365 min
Delta R.T.: 0.016 min
Response: 97808769
Conc: 49.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.697 min
Delta R.T.: -0.013 min
Response: 46038815
Conc: 51.92 ng/ml



#21 AR-1248-1


R.T.: 5.634 min
Delta R.T.: 0.005 min
Response: 23917981
Conc: 437.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.738 min
Delta R.T.: -0.005 min
Response: 12819433
Conc: 449.51 ng/ml




#22 AR-1248-2

R.T.: 5.913 min
Delta R.T.: 0.005 min
Response: 39266997
Conc: 453.96 ng/ml



#22 AR-1248-2


R.T.: 4.976 min
Delta R.T.: -0.005 min
Response: 19375926
Conc: 436.84 ng/ml



#23 AR-1248-3


R.T.: 6.120 min
Delta R.T.: 0.005 min
Response: 42886801
Conc: 466.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.018 min
Delta R.T.: -0.005 min
Response: 20180204
Conc: 442.95 ng/ml




#24 AR-1248-4

R.T.: 6.532 min
Delta R.T.: 0.006 min
Response: 44207981
Conc: 497.01 ng/ml



#24 AR-1248-4


R.T.: 5.189 min
Delta R.T.: -0.005 min
Response: 22823471
Conc: 437.72 ng/ml



#25 AR-1248-5

R.T.: 6.571 min
Delta R.T.: 0.005 min
Response: 42740585
Conc: 494.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.584 min
Delta R.T.: -0.006 min
Response: 20562696
Conc: 464.92 ng/ml