

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011823\
 Data File : P0092171.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jan 2023 07:11
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 19 08:03:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 11 04:41:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.447	3.647	124.9E6	57103666	44.773	41.130
2) SA Decachlor...	10.365	8.692	103.1E6	49693225	52.672	56.043
Target Compounds						
21) L5 AR-1248-1	5.633	4.735	20601500	12393431	376.418	434.577
22) L5 AR-1248-2	5.913	4.973	35847770	18594652	414.434	419.227
23) L5 AR-1248-3	6.120	5.014	40098061	19485514	435.841	427.704
24) L5 AR-1248-4	6.531	5.186	41536850	21960604	466.975	421.169
25) L5 AR-1248-5	6.571	5.580	40105337	19682035	463.592	445.012



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO011823\
Data File : PO092171.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jan 2023 07:11
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 19 08:03:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 11 04:41:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.447 min
Delta R.T.: 0.002 min
Response: 124880432
Conc: 44.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.647 min
Delta R.T.: -0.004 min
Response: 57103666
Conc: 41.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.365 min
Delta R.T.: 0.016 min
Response: 103115080
Conc: 52.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.692 min
Delta R.T.: -0.018 min
Response: 49693225
Conc: 56.04 ng/ml



#21 AR-1248-1


R.T.: 5.633 min
Delta R.T.: 0.004 min
Response: 20601500
Conc: 376.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.735 min
Delta R.T.: -0.008 min
Response: 12393431
Conc: 434.58 ng/ml




#22 AR-1248-2

R.T.: 5.913 min
Delta R.T.: 0.005 min
Response: 35847770
Conc: 414.43 ng/ml



#22 AR-1248-2


R.T.: 4.973 min
Delta R.T.: -0.008 min
Response: 18594652
Conc: 419.23 ng/ml



#23 AR-1248-3


R.T.: 6.120 min
Delta R.T.: 0.004 min
Response: 40098061
Conc: 435.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.014 min
Delta R.T.: -0.008 min
Response: 19485514
Conc: 427.70 ng/ml




#24 AR-1248-4

R.T.: 6.531 min
Delta R.T.: 0.005 min
Response: 41536850
Conc: 466.98 ng/ml



#24 AR-1248-4


R.T.: 5.186 min
Delta R.T.: -0.008 min
Response: 21960604
Conc: 421.17 ng/ml



#25 AR-1248-5

R.T.: 6.571 min
Delta R.T.: 0.006 min
Response: 40105337
Conc: 463.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.580 min
Delta R.T.: -0.009 min
Response: 19682035
Conc: 445.01 ng/ml