

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0011922\
 Data File : P0084346.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Jan 2022 17:50
 Operator : AJ\MA
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1660ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/20/2022
 Supervised By :Ankita Jodhani 01/20/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 02:55:22 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0011922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 20 02:54:17 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	6.520	5.305	5473246	1592979	69.659	70.725
2)	SA Decachlor...	13.794	11.311	6099780	2272479	71.314	71.977
Target Compounds							
3)	L1 AR-1016-1	7.971	6.774	2047899	648827	687.496	695.389
4)	L1 AR-1016-2	7.996	6.797	2954599	930451	683.632	693.265
5)	L1 AR-1016-3	8.069	7.014	1905210	502005	687.091	702.528
6)	L1 AR-1016-4	8.184	7.067	1512980	406690	680.178	693.900
7)	L1 AR-1016-5	8.519	7.324	1532285	529063	676.633	698.915
31)	L7 AR-1260-1	9.760	8.500	2634115	1005134	662.540m	694.357
32)	L7 AR-1260-2	10.051	8.703	3216460	1226599	691.476	702.426
33)	L7 AR-1260-3	10.490	8.871	2565543	1170243	683.886	701.344
34)	L7 AR-1260-4	10.788	9.375	3069602	1038743	691.119	705.163
35)	L7 AR-1260-5	11.212	9.625	5820294	2577643	703.082	717.471

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0011922\
Data File : P0084346.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Jan 2022 17:50
Operator : AJ\MA
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/20/2022
Supervised By :Ankita Jodhani 01/20/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 20 02:55:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011922.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 20 02:54:17 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

