

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012020\
 Data File : P0065664.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jan 2020 15:13
 Operator : DD\AJ
 Sample : L1188-01MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 16:46:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.160	3.434	437651	479351	23.967	21.791
2)	SA Decachlor...	9.663	8.337	748109	905901	22.080	21.845
Target Compounds							
3)	L1 AR-1016-1	5.311	4.511	211749	336016	242.789	326.116 #
4)	L1 AR-1016-2	5.332	4.511	313171	336016	244.153	231.875
5)	L1 AR-1016-3	5.393	4.684	190318	174917	241.558	224.986
6)	L1 AR-1016-4	5.490	4.726	155133	144463	240.433	216.079
7)	L1 AR-1016-5	5.780	4.935	160366	193496	241.900	227.411
8)	L2 AR-1221-1	0.000	3.641	0	16638	N.D.	64.170 #
9)	L2 AR-1221-2	4.455	3.727	23872	28167	152.897	142.853
10)	L2 AR-1221-3	4.528	3.801	91996	104489	176.166	162.175
11)	L3 AR-1232-1	4.528	3.801	91996	104489	227.452	209.731
12)	L3 AR-1232-2	5.046	4.511	135003	336016	641.979	623.126
13)	L3 AR-1232-3	5.332	4.684	313171	174917	670.663	614.899
14)	L3 AR-1232-4	5.490	4.766	155133	176065	664.935	658.598
15)	L3 AR-1232-5	5.580	4.935	132141	193496	747.398	674.945
16)	L4 AR-1242-1	5.311	4.494	211749	232049	363.716	340.443
17)	L4 AR-1242-2	5.332	4.511	313171	336016	366.246	350.627
18)	L4 AR-1242-3	5.393	4.684	190318	174917	362.935	341.258
19)	L4 AR-1242-4	5.490	4.766	155133	176065	366.178	336.860
20)	L4 AR-1242-5	6.181	5.281	23635	161268	43.405	221.888 #
21)	L5 AR-1248-1	5.311	4.511	211749	336016	455.071	606.129 #
22)	L5 AR-1248-2	5.580	4.726	132141	144463	200.258	188.290
23)	L5 AR-1248-3	5.780	4.766	160366	176065	212.832	223.777
24)	L5 AR-1248-4	6.181	4.935	23635	193496	24.951	201.873 #
25)	L5 AR-1248-5	6.181	5.322	23635	22824	25.839	22.547
26)	L6 AR-1254-1	6.153	5.281	123619	161268	145.721	111.318
27)	L6 AR-1254-2	6.369	5.428	149446	164965	108.750	124.715
28)	L6 AR-1254-3	6.732	5.840	96231	306467	62.089	135.995 #
29)	L6 AR-1254-4	7.015	6.053	67755	44196	53.153	29.315 #
30)	L6 AR-1254-5	7.429	6.465	556771	646968	396.470	290.650 #
31)	L7 AR-1260-1	6.894	5.953	355669	443811	257.875	235.737
32)	L7 AR-1260-2	7.149	6.142	467614	561225	266.519	231.005
33)	L7 AR-1260-3	7.505	6.292	320785	498014	226.960	230.082
34)	L7 AR-1260-4	7.727	6.759	369064	378566	211.368	188.695
35)	L7 AR-1260-5	8.034	7.001	736251	927002	219.541	187.255
36)	L8 AR-1262-1	7.505	6.556	320785	205318	193.914	217.787
37)	L8 AR-1262-2	8.034	7.001	736251	927002	229.908	206.307
38)	L8 AR-1262-3	8.332	7.280	479949	187696	210.834	111.522 #
39)	L8 AR-1262-4	8.400	7.343	116156	689642	65.710	200.536 #
40)	L8 AR-1262-5	8.996	7.836	187566	216633	140.999	127.041
41)	L9 AR-1268-1	8.332	7.280	479949	187696	117.975	35.342 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012020\
 Data File : P0065664.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jan 2020 15:13
 Operator : DD\AJ
 Sample : L1188-01MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 16:46:43 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.400	7.343	116156	689642	30.181	135.769 #
43) L9	AR-1268-3	0.000	7.547	0	21402	N.D.	5.038 #
44) L9	AR-1268-4	8.996	7.836	187566	216633	122.108	109.453
45) L9	AR-1268-5	9.365	8.107	75012	74607	6.770	5.189



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012020\
Data File : P0065664.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jan 2020 15:13
Operator : DD\AJ
Sample : L1188-01MSD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 20 16:46:43 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.160 min
Delta R.T.: -0.007 min
Response: 437651
Conc: 23.97 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.434 min
Delta R.T.: -0.007 min
Response: 479351
Conc: 21.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.663 min
Delta R.T.: -0.015 min
Response: 748109
Conc: 22.08 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.337 min
Delta R.T.: -0.016 min
Response: 905901
Conc: 21.84 ng/ml



#3 AR-1016-1


R.T.: 5.311 min
Delta R.T.: -0.008 min
Response: 211749
Conc: 242.79 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.511 min
Delta R.T.: 0.008 min
Response: 336016
Conc: 326.12 ng/ml




#4 AR-1016-2

R.T.: 5.332 min
Delta R.T.: -0.008 min
Response: 313171
Conc: 244.15 ng/ml



#4 AR-1016-2


R.T.: 4.511 min
Delta R.T.: -0.009 min
Response: 336016
Conc: 231.87 ng/ml



#5 AR-1016-3


R.T.: 5.393 min
Delta R.T.: -0.008 min
Response: 190318
Conc: 241.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.684 min
Delta R.T.: -0.010 min
Response: 174917
Conc: 224.99 ng/ml




#6 AR-1016-4

R.T.: 5.490 min
Delta R.T.: -0.008 min
Response: 155133
Conc: 240.43 ng/ml



#6 AR-1016-4


R.T.: 4.726 min
Delta R.T.: -0.009 min
Response: 144463
Conc: 216.08 ng/ml



#7 AR-1016-5


R.T.: 5.780 min
Delta R.T.: -0.008 min
Response: 160366
Conc: 241.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 4.935 min
Delta R.T.: -0.010 min
Response: 193496
Conc: 227.41 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.375 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.641 min
Delta R.T.: -0.009 min
Response: 16638
Conc: 64.17 ng/ml



#9 AR-1221-2


R.T.: 4.455 min
Delta R.T.: -0.005 min
Response: 23872
Conc: 152.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.727 min
Delta R.T.: -0.006 min
Response: 28167
Conc: 142.85 ng/ml




#10 AR-1221-3

R.T.: 4.528 min
Delta R.T.: -0.007 min
Response: 91996
Conc: 176.17 ng/ml



#10 AR-1221-3


R.T.: 3.801 min
Delta R.T.: -0.007 min
Response: 104489
Conc: 162.18 ng/ml



#11 AR-1232-1


R.T.: 4.528 min
Delta R.T.: -0.007 min
Response: 91996
Conc: 227.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.801 min
Delta R.T.: -0.007 min
Response: 104489
Conc: 209.73 ng/ml




#12 AR-1232-2

R.T.: 5.046 min
Delta R.T.: -0.007 min
Response: 135003
Conc: 641.98 ng/ml



#12 AR-1232-2


R.T.: 4.511 min
Delta R.T.: -0.009 min
Response: 336016
Conc: 623.13 ng/ml



#13 AR-1232-3


R.T.: 5.332 min
Delta R.T.: -0.008 min
Response: 313171
Conc: 670.66 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.684 min
Delta R.T.: -0.009 min
Response: 174917
Conc: 614.90 ng/ml




#14 AR-1232-4

R.T.: 5.490 min
Delta R.T.: -0.008 min
Response: 155133
Conc: 664.94 ng/ml



#14 AR-1232-4


R.T.: 4.766 min
Delta R.T.: -0.010 min
Response: 176065
Conc: 658.60 ng/ml



#15 AR-1232-5


R.T.: 5.580 min
Delta R.T.: -0.008 min
Response: 132141
Conc: 747.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 4.935 min
Delta R.T.: -0.009 min
Response: 193496
Conc: 674.94 ng/ml




#16 AR-1242-1

R.T.: 5.311 min
Delta R.T.: -0.008 min
Response: 211749
Conc: 363.72 ng/ml



#16 AR-1242-1


R.T.: 4.494 min
Delta R.T.: -0.008 min
Response: 232049
Conc: 340.44 ng/ml



#17 AR-1242-2


R.T.: 5.332 min
Delta R.T.: -0.008 min
Response: 313171
Conc: 366.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.511 min
Delta R.T.: -0.009 min
Response: 336016
Conc: 350.63 ng/ml




#18 AR-1242-3

R.T.: 5.393 min
Delta R.T.: -0.007 min
Response: 190318
Conc: 362.93 ng/ml



#18 AR-1242-3


R.T.: 4.684 min
Delta R.T.: -0.009 min
Response: 174917
Conc: 341.26 ng/ml



#19 AR-1242-4


R.T.: 5.490 min
Delta R.T.: -0.007 min
Response: 155133
Conc: 366.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.766 min
Delta R.T.: -0.009 min
Response: 176065
Conc: 336.86 ng/ml




#20 AR-1242-5

R.T.: 6.181 min
Delta R.T.: -0.045 min
Response: 23635
Conc: 43.40 ng/ml



#20 AR-1242-5


R.T.: 5.281 min
Delta R.T.: -0.010 min
Response: 161268
Conc: 221.89 ng/ml



#21 AR-1248-1


R.T.: 5.311 min
Delta R.T.: -0.009 min
Response: 211749
Conc: 455.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.511 min
Delta R.T.: 0.008 min
Response: 336016
Conc: 606.13 ng/ml




#22 AR-1248-2

R.T.: 5.580 min
Delta R.T.: -0.009 min
Response: 132141
Conc: 200.26 ng/ml



#22 AR-1248-2


R.T.: 4.726 min
Delta R.T.: -0.010 min
Response: 144463
Conc: 188.29 ng/ml



#23 AR-1248-3


R.T.: 5.780 min
Delta R.T.: -0.009 min
Response: 160366
Conc: 212.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.766 min
Delta R.T.: -0.010 min
Response: 176065
Conc: 223.78 ng/ml




#24 AR-1248-4

R.T.: 6.181 min
Delta R.T.: -0.009 min
Response: 23635
Conc: 24.95 ng/ml



#24 AR-1248-4


R.T.: 4.935 min
Delta R.T.: -0.010 min
Response: 193496
Conc: 201.87 ng/ml



#25 AR-1248-5


R.T.: 6.181 min
Delta R.T.: -0.046 min
Response: 23635
Conc: 25.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.322 min
Delta R.T.: -0.011 min
Response: 22824
Conc: 22.55 ng/ml




#26 AR-1254-1

R.T.: 6.153 min
Delta R.T.: -0.009 min
Response: 123619
Conc: 145.72 ng/ml



#26 AR-1254-1


R.T.: 5.281 min
Delta R.T.: -0.011 min
Response: 161268
Conc: 111.32 ng/ml



#27 AR-1254-2


R.T.: 6.369 min
Delta R.T.: -0.010 min
Response: 149446
Conc: 108.75 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.428 min
Delta R.T.: -0.011 min
Response: 164965
Conc: 124.72 ng/ml




#28 AR-1254-3

R.T.: 6.732 min
Delta R.T.: -0.009 min
Response: 96231
Conc: 62.09 ng/ml



#28 AR-1254-3


R.T.: 5.840 min
Delta R.T.: 0.002 min
Response: 306467
Conc: 135.99 ng/ml



#29 AR-1254-4


R.T.: 7.015 min
Delta R.T.: -0.011 min
Response: 67755
Conc: 53.15 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.053 min
Delta R.T.: -0.011 min
Response: 44196
Conc: 29.32 ng/ml




#30 AR-1254-5

R.T.: 7.429 min
Delta R.T.: -0.011 min
Response: 556771
Conc: 396.47 ng/ml



#30 AR-1254-5


R.T.: 6.465 min
Delta R.T.: -0.013 min
Response: 646968
Conc: 290.65 ng/ml



#31 AR-1260-1


R.T.: 6.894 min
Delta R.T.: -0.009 min
Response: 355669
Conc: 257.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 5.953 min
Delta R.T.: -0.012 min
Response: 443811
Conc: 235.74 ng/ml




#32 AR-1260-2

R.T.: 7.149 min
Delta R.T.: -0.010 min
Response: 467614
Conc: 266.52 ng/ml



#32 AR-1260-2


R.T.: 6.142 min
Delta R.T.: -0.011 min
Response: 561225
Conc: 231.01 ng/ml



#33 AR-1260-3


R.T.: 7.505 min
Delta R.T.: -0.009 min
Response: 320785
Conc: 226.96 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.292 min
Delta R.T.: -0.012 min
Response: 498014
Conc: 230.08 ng/ml




#34 AR-1260-4

R.T.: 7.727 min
Delta R.T.: -0.010 min
Response: 369064
Conc: 211.37 ng/ml



#34 AR-1260-4


R.T.: 6.759 min
Delta R.T.: -0.012 min
Response: 378566
Conc: 188.70 ng/ml



#35 AR-1260-5


R.T.: 8.034 min
Delta R.T.: -0.011 min
Response: 736251
Conc: 219.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.001 min
Delta R.T.: -0.013 min
Response: 927002
Conc: 187.26 ng/ml




#36 AR-1262-1

R.T.: 7.505 min
Delta R.T.: -0.010 min
Response: 320785
Conc: 193.91 ng/ml



#36 AR-1262-1


R.T.: 6.556 min
Delta R.T.: -0.013 min
Response: 205318
Conc: 217.79 ng/ml



#37 AR-1262-2


R.T.: 8.034 min
Delta R.T.: -0.011 min
Response: 736251
Conc: 229.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.001 min
Delta R.T.: -0.013 min
Response: 927002
Conc: 206.31 ng/ml




#38 AR-1262-3

R.T.: 8.332 min
Delta R.T.: 0.002 min
Response: 479949
Conc: 210.83 ng/ml



#38 AR-1262-3


R.T.: 7.280 min
Delta R.T.: -0.013 min
Response: 187696
Conc: 111.52 ng/ml



#39 AR-1262-4


R.T.: 8.400 min
Delta R.T.: -0.013 min
Response: 116156
Conc: 65.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.343 min
Delta R.T.: -0.015 min
Response: 689642
Conc: 200.54 ng/ml




#40 AR-1262-5

R.T.: 8.996 min
Delta R.T.: -0.014 min
Response: 187566
Conc: 141.00 ng/ml



#40 AR-1262-5


R.T.: 7.836 min
Delta R.T.: -0.014 min
Response: 216633
Conc: 127.04 ng/ml



#41 AR-1268-1


R.T.: 8.332 min
Delta R.T.: 0.005 min
Response: 479949
Conc: 117.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.280 min
Delta R.T.: -0.013 min
Response: 187696
Conc: 35.34 ng/ml




#42 AR-1268-2

R.T.: 8.400 min
Delta R.T.: -0.014 min
Response: 116156
Conc: 30.18 ng/ml



#42 AR-1268-2


R.T.: 7.343 min
Delta R.T.: -0.015 min
Response: 689642
Conc: 135.77 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 8.619 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.547 min
Delta R.T.: -0.014 min
Response: 21402
Conc: 5.04 ng/ml




#44 AR-1268-4

R.T.: 8.996 min
Delta R.T.: -0.014 min
Response: 187566
Conc: 122.11 ng/ml



#44 AR-1268-4


R.T.: 7.836 min
Delta R.T.: -0.013 min
Response: 216633
Conc: 109.45 ng/ml



#45 AR-1268-5

R.T.: 9.365 min
Delta R.T.: -0.015 min
Response: 75012
Conc: 6.77 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.107 min
Delta R.T.: -0.014 min
Response: 74607
Conc: 5.19 ng/ml