

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012021\
 Data File : P0074994.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jan 2021 14:59
 Operator : AJ/MA
 Sample : PB134242BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB134242BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 21 05:24:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 06:02:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.928	4.002	2216077	1096743	21.002	19.497
2) SA Decachlor...	10.944	9.252	2104542	1047483	20.736	19.078
Target Compounds						
8) L2 AR-1221-1	5.182	0.000	65092	0	52.994	N.D. #
27) L6 AR-1254-2	7.398	0.000	51783	0	7.603	N.D. #
30) L6 AR-1254-5	8.507	0.000	30674	0	5.281	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012021\
Data File : P0074994.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jan 2021 14:59
Operator : AJ/MA
Sample : PB134242BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB134242BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 21 05:24:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 06:02:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.928 min
Delta R.T.: -0.003 min
Response: 2216077
Conc: 21.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB134242BL




#1 Tetrachloro-m-xylene

R.T.: 4.002 min
Delta R.T.: -0.008 min
Response: 1096743
Conc: 19.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.944 min
Delta R.T.: -0.003 min
Response: 2104542
Conc: 20.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.252 min
Delta R.T.: -0.009 min
Response: 1047483
Conc: 19.08 ng/ml



#8 AR-1221-1


R.T.: 5.182 min
Delta R.T.: 0.000 min
Response: 65092
Conc: 52.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB134242BL




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.263 min
Response: 0
Conc: N.D.




#27 AR-1254-2

R.T.: 7.398 min
Delta R.T.: 0.000 min
Response: 51783
Conc: 7.60 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T. : 6.275 min
Response: 0
Conc: N.D.



#30 AR-1254-5

R.T.: 8.507 min
Delta R.T.: 0.000 min
Response: 30674
Conc: 5.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB134242BL



#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.356 min
Response: 0
Conc: N.D.