

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012220\
 Data File : P0065791.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jan 2020 18:56
 Operator : DD\AJ
 Sample : L1161-03
 Misc : AR1221 MDL 25 PPB
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 01:12:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.161	3.436	346587	430239	18.980	19.559
2)	SA Decachlor...	9.658	8.334	626239	749795	18.483	18.080
Target Compounds							
3)	L1 AR-1016-1	5.247	0.000	24186	0	27.731	N.D. #
8)	L2 AR-1221-1	4.371	0.000	5721	0	27.599	N.D. #
9)	L2 AR-1221-2	4.458	3.728	9350	9619	59.885	48.784
10)	L2 AR-1221-3	4.529	3.802	15796	19395	30.249	30.102
11)	L3 AR-1232-1	4.529	3.802	15796	19395	39.055	38.929
16)	L4 AR-1242-1	5.247	0.000	24186	0	41.544	N.D. #
28)	L6 AR-1254-3	6.740	0.000	12439	0	8.026	N.D. #
29)	L6 AR-1254-4	0.000	6.129	0	12165	N.D.	8.069 #
32)	L7 AR-1260-2	0.000	6.129	0	12165	N.D.	5.007 #
43)	L9 AR-1268-3	8.599	0.000	5954	0	1.850	N.D. #
45)	L9 AR-1268-5	9.362	0.000	8828	0	0.797	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012220\
Data File : P0065791.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jan 2020 18:56
Operator : DD\AJ
Sample : L1161-03
Misc : AR1221 MDL 25 PPB
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 01:12:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.161 min
Delta R.T.: -0.006 min
Response: 346587
Conc: 18.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.436 min
Delta R.T.: -0.005 min
Response: 430239
Conc: 19.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.658 min
Delta R.T.: -0.020 min
Response: 626239
Conc: 18.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.334 min
Delta R.T.: -0.019 min
Response: 749795
Conc: 18.08 ng/ml



#3 AR-1016-1


R.T.: 5.247 min
Delta R.T.: -0.063 min
Response: 24186
Conc: 27.73 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.490 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.371 min
Delta R.T.: -0.004 min
Response: 5721
Conc: 27.60 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.650 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.458 min
Delta R.T.: -0.003 min
Response: 9350
Conc: 59.89 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.728 min
Delta R.T.: -0.006 min
Response: 9619
Conc: 48.78 ng/ml




#10 AR-1221-3

R.T.: 4.529 min
Delta R.T.: -0.007 min
Response: 15796
Conc: 30.25 ng/ml



#10 AR-1221-3


R.T.: 3.802 min
Delta R.T.: -0.006 min
Response: 19395
Conc: 30.10 ng/ml



#11 AR-1232-1


R.T.: 4.529 min
Delta R.T.: -0.006 min
Response: 15796
Conc: 39.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.802 min
Delta R.T.: -0.006 min
Response: 19395
Conc: 38.93 ng/ml




#16 AR-1242-1

R.T.: 5.247 min
Delta R.T.: -0.063 min
Response: 24186
Conc: 41.54 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T.: 4.490 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 6.740 min
Delta R.T.: -0.001 min
Response: 12439
Conc: 8.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 5.837 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 7.026 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.129 min
Delta R.T.: 0.065 min
Response: 12165
Conc: 8.07 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 7.160 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.129 min
Delta R.T.: -0.024 min
Response: 12165
Conc: 5.01 ng/ml




#43 AR-1268-3

R.T.: 8.599 min
Delta R.T.: -0.019 min
Response: 5954
Conc: 1.85 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 7.560 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.362 min
Delta R.T.: -0.018 min
Response: 8828
Conc: 0.80 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.120 min
Response: 0
Conc: N.D.