

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012220\
 Data File : P0065782.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jan 2020 16:26
 Operator : DD\AJ
 Sample : L1161-08
 Misc : AR1254 LOQ 50 PPB
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOQ-WATER-02-QT1-2020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 01:25:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 04 04:01:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.162	3.437	224613	285088	12.301	12.960
2)	SA Decachlor...	9.659	8.335	437551	535508	12.914	12.913
Target Compounds							
26)	L6 AR-1254-1	6.153	5.283	56708	81311	66.846	56.126
27)	L6 AR-1254-2	6.369	5.430	70161	89056	51.055	67.327 #
28)	L6 AR-1254-3	6.733	5.826	87172	106762	56.244	47.376
29)	L6 AR-1254-4	7.015	6.053	65192	72030	51.143	47.778
30)	L6 AR-1254-5	7.429	6.466	59111	100341	42.092	45.078



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012220\
Data File : P0065782.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jan 2020 16:26
Operator : DD\AJ
Sample : L1161-08
Misc : AR1254 LOQ 50 PPB
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2020

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 01:25:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010420.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 04 04:01:28 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.162 min
Delta R.T.: -0.005 min
Response: 224613
Conc: 12.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2020




#1 Tetrachloro-m-xylene

R.T.: 3.437 min
Delta R.T.: -0.004 min
Response: 285088
Conc: 12.96 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.659 min
Delta R.T.: -0.019 min
Response: 437551
Conc: 12.91 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.335 min
Delta R.T.: -0.017 min
Response: 535508
Conc: 12.91 ng/ml



#26 AR-1254-1


R.T.: 6.153 min
Delta R.T.: -0.009 min
Response: 56708
Conc: 66.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2020




#26 AR-1254-1

R.T.: 5.283 min
Delta R.T.: -0.010 min
Response: 81311
Conc: 56.13 ng/ml




#27 AR-1254-2

R.T.: 6.369 min
Delta R.T.: -0.009 min
Response: 70161
Conc: 51.06 ng/ml



#27 AR-1254-2


R.T.: 5.430 min
Delta R.T.: -0.009 min
Response: 89056
Conc: 67.33 ng/ml



#28 AR-1254-3


R.T.: 6.733 min
Delta R.T.: -0.009 min
Response: 87172
Conc: 56.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2020




#28 AR-1254-3

R.T.: 5.826 min
Delta R.T.: -0.011 min
Response: 106762
Conc: 47.38 ng/ml




#29 AR-1254-4

R.T.: 7.015 min
Delta R.T.: -0.011 min
Response: 65192
Conc: 51.14 ng/ml



#29 AR-1254-4


R.T.: 6.053 min
Delta R.T.: -0.011 min
Response: 72030
Conc: 47.78 ng/ml



#30 AR-1254-5

R.T.: 7.429 min
Delta R.T.: -0.012 min
Response: 59111
Conc: 42.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2020



#30 AR-1254-5

R.T.: 6.466 min
Delta R.T.: -0.012 min
Response: 100341
Conc: 45.08 ng/ml