

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012224\
 Data File : P0101223.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jan 2024 15:34
 Operator : YP/AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 17:20:15 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jan 22 17:18:40 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.333	3.631	149.9E6	105.9E6	50.000	50.000
2)	SA Decachlor...	10.062	8.677	77169083	69079944	50.000	50.000
Target Compounds							
26)	L6 AR-1254-1	6.358	5.532	44535758	48473488	500.000	500.000
27)	L6 AR-1254-2	6.576	5.681	62728278	42461660	500.000	500.000
28)	L6 AR-1254-3	6.944	6.088	58025868	62872819	500.000	500.000
29)	L6 AR-1254-4	7.230	6.318	34913178	30935128	500.000	500.000
30)	L6 AR-1254-5	7.650	6.738	41867013	51066863	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012224\
Data File : PO101223.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jan 2024 15:34
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 17:20:15 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012224.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jan 22 17:18:40 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

