

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012224\
 Data File : P0101235.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jan 2024 18:54
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO012224AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 21:33:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jan 22 18:08:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.334	3.632	137.9E6	97403453	48.849	50.199
2) SA Decachlor...	10.062	8.678	69941074	62965105	49.763	49.586
Target Compounds						
26) L6 AR-1254-1	6.358	5.533	40994378	44704612	468.279	462.250
27) L6 AR-1254-2	6.577	5.681	57802117	39457672	461.912	463.857
28) L6 AR-1254-3	6.944	6.089	53352520	57847717	466.573	468.089
29) L6 AR-1254-4	7.231	6.319	32228220	28714498	463.665	471.145
30) L6 AR-1254-5	7.652	6.739	38169583	46574240	461.584	460.300

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012224\
Data File : PO101235.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jan 2024 18:54
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO012224AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 21:33:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012224.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jan 22 18:08:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

