

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012320\
 Data File : P0065862.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Jan 2020 21:33
 Operator : DD\AJ
 Sample : P0012320ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO012320

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 04:00:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 24 03:54:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	23.188	22.350 E3	3.6	94	0.00
2 SA Decachlorobiphenyl	41.627	40.650 E3	2.3	99	0.00
3 L1 AR-1016-1	1.098	1.048 E3	4.6	97	0.00
4 L1 AR-1016-2	1.654	1.571 E3	5.0	96	0.00
5 L1 AR-1016-3	1.018	0.969 E3	4.8	97	0.00
6 L1 AR-1016-4	838.526	802.589	4.3	96	0.00
7 L1 AR-1016-5	854.153	830.090	2.8	97	0.00
31 L7 AR-1260-1	1.938	1.846 E3	4.7	99	0.00
32 L7 AR-1260-2	2.622	2.522 E3	3.8	100	0.00
33 L7 AR-1260-3	2.173	2.141 E3	1.5	101	0.00
34 L7 AR-1260-4	2.217	2.190 E3	1.2	102	0.00
35 L7 AR-1260-5	4.842	4.764 E3	1.6	101	0.00

Signal #2

1 SA Tetrachloro-m-xylene	26.561	26.936 E3	-1.4	100	0.00
2 SA Decachlorobiphenyl	48.596	48.407 E3	0.4	101	0.00
3 L1 AR-1016-1	1.281	1.237 E3	3.4	102	0.00
4 L1 AR-1016-2	1.880	1.842 E3	2.0	101	0.00
5 L1 AR-1016-3	970.357	967.362	0.3	101	0.00
6 L1 AR-1016-4	820.882	827.264	-0.8	103	0.00
7 L1 AR-1016-5	1.078	1.064 E3	1.3	101	0.00
31 L7 AR-1260-1	2.422	2.361 E3	2.5	102	0.00
32 L7 AR-1260-2	3.169	3.072 E3	3.1	103	0.00
33 L7 AR-1260-3	2.765	2.728 E3	1.3	103	0.00
34 L7 AR-1260-4	2.468	2.454 E3	0.6	103	0.00
35 L7 AR-1260-5	6.079	6.079 E3	0.0	103	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	289.071	0.000	100.0#	0#	-4.36#
9 L2 AR-1221-2	220.581	0.000	100.0#	0#	-4.45#
10 L2 AR-1221-3	717.417	0.000	100.0#	0#	-4.52#
11 L3 AR-1232-1	543.076	0.000	100.0#	0#	-4.52#
12 L3 AR-1232-2	288.819	0.000	100.0#	0#	-5.04#
13 L3 AR-1232-3	673.575	0.000	100.0#	0#	-5.33#
14 L3 AR-1232-4	336.045	0.000	100.0#	0#	-5.48#
15 L3 AR-1232-5	248.045	0.000	100.0#	0#	-5.57#
16 L4 AR-1242-1	805.902	0.000	100.0#	0#	-5.30#
17 L4 AR-1242-2	1.196	0.000 E3	100.0#	0#	-5.33#
18 L4 AR-1242-3	749.877	0.000	100.0#	0#	-5.39#
19 L4 AR-1242-4	611.733	0.000	100.0#	0#	-5.48#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012320\
 Data File : P0065862.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Jan 2020 21:33
 Operator : DD\AJ
 Sample : P0012320ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO012320

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 04:00:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 24 03:54:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
20 L4 AR-1242-5	832.672	0.000	100.0#	0#	-6.21#
21 L5 AR-1248-1	617.546	0.000	100.0#	0#	-5.30#
22 L5 AR-1248-2	880.465	0.000	100.0#	0#	-5.57#
23 L5 AR-1248-3	1.032	0.000 E3	100.0#	0#	-5.77#
24 L5 AR-1248-4	1.347	0.000 E3	100.0#	0#	-6.17#
25 L5 AR-1248-5	1.292	0.000 E3	100.0#	0#	-6.21#
26 L6 AR-1254-1	1.327	0.000 E3	100.0#	0#	-6.15#
27 L6 AR-1254-2	2.178	0.000 E3	100.0#	0#	-6.36#
28 L6 AR-1254-3	2.432	0.000 E3	100.0#	0#	-6.72#
29 L6 AR-1254-4	2.096	0.000 E3	100.0#	0#	-7.01#
30 L6 AR-1254-5	2.146	0.000 E3	100.0#	0#	-7.42#
36 L8 AR-1262-1	2.972	0.000 E3	100.0#	0#	-7.50#
37 L8 AR-1262-2	5.149	0.000 E3	100.0#	0#	-8.02#
38 L8 AR-1262-3	3.511	0.000 E3	100.0#	0#	-8.31#
39 L8 AR-1262-4	2.734	0.000 E3	100.0#	0#	-8.39#
40 L8 AR-1262-5	1.886	0.000 E3	100.0#	0#	-8.98#
41 L9 AR-1268-1	6.555	0.000 E3	100.0#	0#	-8.31#
42 L9 AR-1268-2	5.900	0.000 E3	100.0#	0#	-8.39#
43 L9 AR-1268-3	5.039	0.000 E3	100.0#	0#	-8.60#
44 L9 AR-1268-4	2.119	0.000 E3	100.0#	0#	-8.98#
45 L9 AR-1268-5	14.986	0.000 E3	100.0#	0#	-9.35#

Signal #2

8 L2 AR-1221-1	344.275	0.000	100.0#	0#	-3.64#
9 L2 AR-1221-2	254.114	0.000	100.0#	0#	-3.72#
10 L2 AR-1221-3	822.959	0.000	100.0#	0#	-3.80#
11 L3 AR-1232-1	636.523	0.000	100.0#	0#	-3.80#
12 L3 AR-1232-2	768.633	0.000	100.0#	0#	-4.51#
13 L3 AR-1232-3	384.772	0.000	100.0#	0#	-4.68#
14 L3 AR-1232-4	352.135	0.000	100.0#	0#	-4.76#
15 L3 AR-1232-5	398.478	0.000	100.0#	0#	-4.93#
16 L4 AR-1242-1	928.199	0.000	100.0#	0#	-4.49#
17 L4 AR-1242-2	1.383	0.000 E3	100.0#	0#	-4.51#
18 L4 AR-1242-3	735.959	0.000	100.0#	0#	-4.68#
19 L4 AR-1242-4	744.516	0.000	100.0#	0#	-4.76#
20 L4 AR-1242-5	1.107	0.000 E3	100.0#	0#	-5.28#
21 L5 AR-1248-1	713.434	0.000	100.0#	0#	-4.49#
22 L5 AR-1248-2	1.014	0.000 E3	100.0#	0#	-4.72#
23 L5 AR-1248-3	1.028	0.000 E3	100.0#	0#	-4.76#
24 L5 AR-1248-4	1.276	0.000 E3	100.0#	0#	-4.93#
25 L5 AR-1248-5	1.396	0.000 E3	100.0#	0#	-5.32#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012320\
 Data File : P0065862.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Jan 2020 21:33
 Operator : DD\AJ
 Sample : P0012320ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO012320

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 04:00:48 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 24 03:54:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
26 L6 AR-1254-1	2.236	0.000 E3	100.0#	0#	-5.28#
27 L6 AR-1254-2	2.054	0.000 E3	100.0#	0#	-5.42#
28 L6 AR-1254-3	3.391	0.000 E3	100.0#	0#	-5.82#
29 L6 AR-1254-4	2.291	0.000 E3	100.0#	0#	-6.05#
30 L6 AR-1254-5	3.234	0.000 E3	100.0#	0#	-6.46#
36 L8 AR-1262-1	1.384	0.000 E3	100.0#	0#	-6.55#
37 L8 AR-1262-2	6.323	0.000 E3	100.0#	0#	-7.00#
38 L8 AR-1262-3	2.561	0.000 E3	100.0#	0#	-7.27#
39 L8 AR-1262-4	4.703	0.000 E3	100.0#	0#	-7.34#
40 L8 AR-1262-5	2.103	0.000 E3	100.0#	0#	-7.83#
41 L9 AR-1268-1	7.791	0.000 E3	100.0#	0#	-7.27#
42 L9 AR-1268-2	7.243	0.000 E3	100.0#	0#	-7.34#
43 L9 AR-1268-3	5.986	0.000 E3	100.0#	0#	-7.54#
44 L9 AR-1268-4	2.439	0.000 E3	100.0#	0#	-7.83#
45 L9 AR-1268-5	18.065	0.000 E3	100.0#	0#	-8.10#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0