

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012423\
 Data File : P0092259.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2023 08:12
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 16:06:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.448	3.643	160.0E6	76477438	55.264	54.465
2)	SA Decachlor...	10.373	8.687	112.7E6	60702674	54.583	53.771
Target Compounds							
16)	L4 AR-1242-1	5.636	4.732	42066575	20440758	539.393	524.083
17)	L4 AR-1242-2	5.658	4.750	60626417	28744938	538.593	529.262
18)	L4 AR-1242-3	5.721	4.927	38888832	15398166	535.712	528.708
19)	L4 AR-1242-4	5.820	5.011	30805805	16920898	556.317	526.050
20)	L4 AR-1242-5	6.573	5.536	31034890	20843444	555.961	575.758



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012423\
Data File : PO092259.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2023 08:12
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 16:06:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.448 min
Delta R.T.: -0.001 min
Response: 159982965
Conc: 55.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.643 min
Delta R.T.: -0.003 min
Response: 76477438
Conc: 54.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.373 min
Delta R.T.: 0.002 min
Response: 112656683
Conc: 54.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.687 min
Delta R.T.: -0.006 min
Response: 60702674
Conc: 53.77 ng/ml



#16 AR-1242-1


R.T.: 5.636 min
Delta R.T.: 0.002 min
Response: 42066575
Conc: 539.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.732 min
Delta R.T.: -0.002 min
Response: 20440758
Conc: 524.08 ng/ml




#17 AR-1242-2

R.T.: 5.658 min
Delta R.T.: 0.001 min
Response: 60626417
Conc: 538.59 ng/ml



#17 AR-1242-2


R.T.: 4.750 min
Delta R.T.: -0.002 min
Response: 28744938
Conc: 529.26 ng/ml



#18 AR-1242-3


R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 38888832
Conc: 535.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.927 min
Delta R.T.: -0.002 min
Response: 15398166
Conc: 528.71 ng/ml




#19 AR-1242-4

R.T.: 5.820 min
Delta R.T.: 0.001 min
Response: 30805805
Conc: 556.32 ng/ml



#19 AR-1242-4


R.T.: 5.011 min
Delta R.T.: -0.002 min
Response: 16920898
Conc: 526.05 ng/ml



#20 AR-1242-5

R.T.: 6.573 min
Delta R.T.: 0.001 min
Response: 31034890
Conc: 555.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.536 min
Delta R.T.: -0.003 min
Response: 20843444
Conc: 575.76 ng/ml