

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012423\
 Data File : P0092291.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2023 18:35
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 01:36:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 24 04:40:59 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.448	3.646	141.9E6	67709910	49.020	48.221
2)	SA Decachlor...	10.364	8.688	103.1E6	55722160	49.977	49.359
Target Compounds							
26)	L6 AR-1254-1	6.505	5.538	53453452	37013849	482.800	472.178
27)	L6 AR-1254-2	6.729	5.687	78923869	32830367	485.523	474.490
28)	L6 AR-1254-3	7.101	6.092	80244268	50591982	486.537	473.996
29)	L6 AR-1254-4	7.394	6.321	53331335	28494704	487.555	483.827
30)	L6 AR-1254-5	7.822	6.742	62479695	43892210	444.066	482.249



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012423\
Data File : PO092291.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2023 18:35
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 01:36:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 24 04:40:59 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.448 min
Delta R.T.: -0.001 min
Response: 141906718
Conc: 49.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.646 min
Delta R.T.: 0.000 min
Response: 67709910
Conc: 48.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.364 min
Delta R.T.: -0.007 min
Response: 103148698
Conc: 49.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.688 min
Delta R.T.: -0.005 min
Response: 55722160
Conc: 49.36 ng/ml



#26 AR-1254-1


R.T.: 6.505 min
Delta R.T.: -0.001 min
Response: 53453452
Conc: 482.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.538 min
Delta R.T.: 0.000 min
Response: 37013849
Conc: 472.18 ng/ml




#27 AR-1254-2

R.T.: 6.729 min
Delta R.T.: -0.002 min
Response: 78923869
Conc: 485.52 ng/ml



#27 AR-1254-2


R.T.: 5.687 min
Delta R.T.: 0.000 min
Response: 32830367
Conc: 474.49 ng/ml



#28 AR-1254-3


R.T.: 7.101 min
Delta R.T.: -0.002 min
Response: 80244268
Conc: 486.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.092 min
Delta R.T.: 0.000 min
Response: 50591982
Conc: 474.00 ng/ml




#29 AR-1254-4

R.T.: 7.394 min
Delta R.T.: -0.002 min
Response: 53331335
Conc: 487.55 ng/ml



#29 AR-1254-4


R.T.: 6.321 min
Delta R.T.: -0.001 min
Response: 28494704
Conc: 483.83 ng/ml



#30 AR-1254-5

R.T.: 7.822 min
Delta R.T.: -0.002 min
Response: 62479695
Conc: 444.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.742 min
Delta R.T.: -0.002 min
Response: 43892210
Conc: 482.25 ng/ml