

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012424\
 Data File : P0101268.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jan 2024 19:29
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 00:51:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 00:51:26 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.333	3.631	385.8E6	259.6E6	103.497	102.792
2) SA Decachlor...	10.057	8.676	232.3E6	185.0E6	100.642	99.476
Target Compounds						
16) L4 AR-1242-1	5.500	4.723	76053997	55930796	981.898	969.098
17) L4 AR-1242-2	5.522	4.742	107.5E6	76031754	990.504	989.185
18) L4 AR-1242-3	5.584	4.921	71176981	42242756	987.231	972.237
19) L4 AR-1242-4	5.683	5.004	57719132	47434679	996.891	958.435
20) L4 AR-1242-5	6.420	5.531	52676893	52552882	1007.526	990.965



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012424\
Data File : PO101268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jan 2024 19:29
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 00:51:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 00:51:26 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.333 min
Delta R.T.: 0.000 min
Response: 385788752
Conc: 103.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 259631787
Conc: 102.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.057 min
Delta R.T.: -0.002 min
Response: 232335609
Conc: 100.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: 0.000 min
Response: 184952312
Conc: 99.48 ng/ml



#16 AR-1242-1


R.T.: 5.500 min
Delta R.T.: 0.000 min
Response: 76053997
Conc: 981.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.723 min
Delta R.T.: 0.000 min
Response: 55930796
Conc: 969.10 ng/ml




#17 AR-1242-2

R.T.: 5.522 min
Delta R.T.: 0.000 min
Response: 107507103
Conc: 990.50 ng/ml



#17 AR-1242-2


R.T.: 4.742 min
Delta R.T.: 0.000 min
Response: 76031754
Conc: 989.18 ng/ml



#18 AR-1242-3


R.T.: 5.584 min
Delta R.T.: 0.000 min
Response: 71176981
Conc: 987.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 42242756
Conc: 972.24 ng/ml




#19 AR-1242-4

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 57719132
Conc: 996.89 ng/ml



#19 AR-1242-4


R.T.: 5.004 min
Delta R.T.: 0.000 min
Response: 47434679
Conc: 958.43 ng/ml



#20 AR-1242-5

R.T.: 6.420 min
Delta R.T.: -0.001 min
Response: 52676893
Conc: 1007.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.531 min
Delta R.T.: -0.001 min
Response: 52552882
Conc: 990.96 ng/ml