

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012521\
 Data File : P0075168.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Jan 2021 00:24
 Operator : AJ/MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 1/26/2021 10:16:54 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 26 01:07:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 06:02:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.924	3.997	5955253	2742516	56.438	48.754
2)	SA Decachlor...	10.939	9.248	5480590	2703524	54.000	49.240m
Target Compounds							
3)	L1 AR-1016-1	6.248	5.240	2260336	1104335	552.848	482.817m
4)	L1 AR-1016-2	6.272	5.261	3180652	1482323	558.504	469.122
5)	L1 AR-1016-3	6.338	5.450	1907210	785950	562.018	457.867
6)	L1 AR-1016-4	6.445	5.503	1599530	659772	562.565	469.449
7)	L1 AR-1016-5	6.762	5.729	1530364	776336	559.542	451.466m
31)	L7 AR-1260-1	7.941	6.817	2896936	1524020	580.695	477.707
32)	L7 AR-1260-2	8.207	7.014	3953871	1896400	581.272	446.102
33)	L7 AR-1260-3	8.574	7.166	3317579	1670103	578.809	463.471
34)	L7 AR-1260-4	8.805	7.645	3148749	1464935	575.502	467.833
35)	L7 AR-1260-5	9.131	7.892	6950919	3658189	556.297	460.227

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012521\
Data File : PO075168.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Jan 2021 00:24
Operator : AJ/MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
1/26/2021 10:16:54 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 26 01:07:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 06:02:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.924 min
Delta R.T.: -0.007 min
Response: 5955253
Conc: 56.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
1/26/2021 10:16:54 AM




#1 Tetrachloro-m-xylene

R.T.: 3.997 min
Delta R.T.: -0.013 min
Response: 2742516
Conc: 48.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.939 min
Delta R.T.: -0.007 min
Response: 5480590
Conc: 54.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.248 min
Delta R.T.: -0.014 min
Response: 2703524
Conc: 49.24 ng/ml m




#3 AR-1016-1

R.T.: 6.248 min
Delta R.T.: -0.006 min
Response: 2260336
Conc: 552.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
1/26/2021 10:16:54 AM




#3 AR-1016-1

R.T.: 5.240 min
Delta R.T.: -0.013 min
Response: 1104335
Conc: 482.82 ng/ml m




#4 AR-1016-2

R.T.: 6.272 min
Delta R.T.: -0.006 min
Response: 3180652
Conc: 558.50 ng/ml



#4 AR-1016-2

R.T.: 5.261 min
Delta R.T.: -0.012 min
Response: 1482323
Conc: 469.12 ng/ml




#5 AR-1016-3

R.T.: 6.338 min
Delta R.T.: -0.006 min
Response: 1907210
Conc: 562.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
1/26/2021 10:16:54 AM




#5 AR-1016-3

R.T.: 5.450 min
Delta R.T.: -0.012 min
Response: 785950
Conc: 457.87 ng/ml




#6 AR-1016-4

R.T.: 6.445 min
Delta R.T.: -0.007 min
Response: 1599530
Conc: 562.56 ng/ml



#6 AR-1016-4

R.T.: 5.503 min
Delta R.T.: -0.012 min
Response: 659772
Conc: 469.45 ng/ml




#7 AR-1016-5

R.T.: 6.762 min
Delta R.T.: -0.005 min
Response: 1530364
Conc: 559.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
1/26/2021 10:16:54 AM




#7 AR-1016-5

R.T.: 5.729 min
Delta R.T.: -0.013 min
Response: 776336
Conc: 451.47 ng/ml m




#31 AR-1260-1

R.T.: 7.941 min
Delta R.T.: -0.004 min
Response: 2896936
Conc: 580.70 ng/ml



#31 AR-1260-1

R.T.: 6.817 min
Delta R.T.: -0.012 min
Response: 1524020
Conc: 477.71 ng/ml




#32 AR-1260-2

R.T.: 8.207 min
Delta R.T.: -0.004 min
Response: 3953871
Conc: 581.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
1/26/2021 10:16:54 AM




#32 AR-1260-2

R.T.: 7.014 min
Delta R.T.: -0.011 min
Response: 1896400
Conc: 446.10 ng/ml




#33 AR-1260-3

R.T.: 8.574 min
Delta R.T.: -0.004 min
Response: 3317579
Conc: 578.81 ng/ml



#33 AR-1260-3

R.T.: 7.166 min
Delta R.T.: -0.012 min
Response: 1670103
Conc: 463.47 ng/ml




#34 AR-1260-4

R.T.: 8.805 min
Delta R.T.: -0.004 min
Response: 3148749
Conc: 575.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
1/26/2021 10:16:54 AM




#34 AR-1260-4

R.T.: 7.645 min
Delta R.T.: -0.012 min
Response: 1464935
Conc: 467.83 ng/ml



#35 AR-1260-5

R.T.: 9.131 min
Delta R.T.: -0.005 min
Response: 6950919
Conc: 556.30 ng/ml



#35 AR-1260-5

R.T.: 7.892 min
Delta R.T.: -0.011 min
Response: 3658189
Conc: 460.23 ng/ml