

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012524\
 Data File : P0101296.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2024 07:13
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 11:40:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.334	3.632	194.6E6	130.4E6	50.702	50.461
2) SA Decachlor...	10.058	8.675	119.3E6	96515395	51.207	50.932
Target Compounds						
16) L4 AR-1242-1	5.500	4.724	40867237	30200238	528.170	513.303
17) L4 AR-1242-2	5.522	4.743	57306047	39935178	529.163	514.996
18) L4 AR-1242-3	5.584	4.921	38172443	22572522	537.033	520.715
19) L4 AR-1242-4	5.683	5.004	30614613	25704163	537.511	516.410
20) L4 AR-1242-5	6.421	5.531	28387376	28143605	528.985	531.283



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012524\
Data File : PO101296.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2024 07:13
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 11:40:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.334 min
Delta R.T.: 0.000 min
Response: 194621625
Conc: 50.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.632 min
Delta R.T.: 0.000 min
Response: 130408348
Conc: 50.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.058 min
Delta R.T.: 0.000 min
Response: 119317422
Conc: 51.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.675 min
Delta R.T.: 0.000 min
Response: 96515395
Conc: 50.93 ng/ml



#16 AR-1242-1


R.T.: 5.500 min
Delta R.T.: 0.000 min
Response: 40867237
Conc: 528.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 30200238
Conc: 513.30 ng/ml




#17 AR-1242-2

R.T.: 5.522 min
Delta R.T.: 0.000 min
Response: 57306047
Conc: 529.16 ng/ml



#17 AR-1242-2


R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 39935178
Conc: 515.00 ng/ml



#18 AR-1242-3


R.T.: 5.584 min
Delta R.T.: 0.000 min
Response: 38172443
Conc: 537.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 22572522
Conc: 520.72 ng/ml




#19 AR-1242-4

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 30614613
Conc: 537.51 ng/ml



#19 AR-1242-4


R.T.: 5.004 min
Delta R.T.: 0.000 min
Response: 25704163
Conc: 516.41 ng/ml



#20 AR-1242-5

R.T.: 6.421 min
Delta R.T.: 0.000 min
Response: 28387376
Conc: 528.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.531 min
Delta R.T.: 0.000 min
Response: 28143605
Conc: 531.28 ng/ml