

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0012621\
 Data File : P0075235.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Jan 2021 5:41
 Operator : AJ/MA
 Sample : M1247-13MS
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 2021-01-25-FD01MS

Manual Integrations
 APPROVED

mohammad
 1/27/2021 5:29:21 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 07:31:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0011121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 06:02:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.925	3.997	2339406	1098752	22.171	19.533
2)	SA Decachlor...	10.941	9.248	2552343	1229018	25.148	22.385
Target Compounds							
3)	L1 AR-1016-1	6.248	5.240	2132780	1034451	521.649	452.264m
4)	L1 AR-1016-2	6.272	5.260	2955249	1430350	518.925	452.674
5)	L1 AR-1016-3	6.338	5.450	1766185	754098	520.460	439.311
6)	L1 AR-1016-4	6.446	5.502	1487728	621172	523.243	441.984
7)	L1 AR-1016-5	6.762	5.728	1356896	792717	496.118	460.992m
31)	L7 AR-1260-1	7.941	6.816	3058250	1636079	613.031	512.832
32)	L7 AR-1260-2	8.208	7.013	4249840	2028182	624.783	477.101
33)	L7 AR-1260-3	8.574	7.166	5368217	1912594	936.579	530.764 #
34)	L7 AR-1260-4	8.808	7.645	4383217	1961756	801.127	626.495
35)	L7 AR-1260-5	9.133	7.892	8843730	4412119	707.783	555.077
36)	L8 AR-1262-1	8.574	7.344	5368217	2399574	602.888	943.014 #
37)	L8 AR-1262-2	9.133	7.892	8843730	4412119	577.634	479.873
38)	L8 AR-1262-3	9.441	8.176	10502951	3710332	1048.639	938.299
39)	L8 AR-1262-4	9.537	8.240	8755522	5369049	1175.657	810.397 #
40)	L8 AR-1262-5	10.199	8.734	5385773	2609415	1029.473	845.176

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012621\
Data File : P0075235.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Jan 2021 5:41
Operator : AJ/MA
Sample : M1247-13MS
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
2021-01-25-FD01MS

Manual Integrations
APPROVED

mohammad
1/27/2021 5:29:21 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 07:31:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 06:02:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

