

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012621\
 Data File : P0075172.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Jan 2021 9:44
 Operator : AJ/MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Ankita
 1/28/2021 2:54:24 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 00:53:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 12 06:02:45 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.924	3.996	5581409	2602939	52.895	46.273
2) SA Decachlor...	10.939	9.247	5111913	2596878	50.367	47.298m
Target Compounds						
26) L6 AR-1254-1	7.160	6.104	2244711	1541682	510.377	426.822
27) L6 AR-1254-2	7.391	6.264	3471974	1330146	509.777	426.802
28) L6 AR-1254-3	7.773	6.680	3535456	2120182	504.033	432.344
29) L6 AR-1254-4	8.071	6.918	3092790	1452082	509.968	418.735
30) L6 AR-1254-5	8.500	7.343	2974844	1948525	512.115	430.524

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012621\
Data File : P0075172.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Jan 2021 9:44
Operator : AJ/MA
Sample : AR1254CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
1/28/2021 2:54:24 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 00:53:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0011121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 12 06:02:45 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.924 min
Delta R.T.: -0.007 min
Response: 5581409
Conc: 52.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
1/28/2021 2:54:24 PM




#1 Tetrachloro-m-xylene

R.T.: 3.996 min
Delta R.T.: -0.014 min
Response: 2602939
Conc: 46.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.939 min
Delta R.T.: -0.007 min
Response: 5111913
Conc: 50.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.247 min
Delta R.T.: -0.014 min
Response: 2596878
Conc: 47.30 ng/ml m




#26 AR-1254-1

R.T.: 7.160 min
Delta R.T.: -0.007 min
Response: 2244711
Conc: 510.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
1/28/2021 2:54:24 PM




#26 AR-1254-1

R.T.: 6.104 min
Delta R.T.: -0.012 min
Response: 1541682
Conc: 426.82 ng/ml




#27 AR-1254-2

R.T.: 7.391 min
Delta R.T.: -0.006 min
Response: 3471974
Conc: 509.78 ng/ml



#27 AR-1254-2

R.T.: 6.264 min
Delta R.T.: -0.012 min
Response: 1330146
Conc: 426.80 ng/ml




#28 AR-1254-3

R.T.: 7.773 min
Delta R.T.: -0.008 min
Response: 3535456
Conc: 504.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Ankita
1/28/2021 2:54:24 PM




#28 AR-1254-3

R.T.: 6.680 min
Delta R.T.: -0.012 min
Response: 2120182
Conc: 432.34 ng/ml




#29 AR-1254-4

R.T.: 8.071 min
Delta R.T.: -0.006 min
Response: 3092790
Conc: 509.97 ng/ml



#29 AR-1254-4

R.T.: 6.918 min
Delta R.T.: -0.012 min
Response: 1452082
Conc: 418.74 ng/ml




#30 AR-1254-5

R.T.: 8.500 min
Delta R.T.: -0.007 min
Response: 2974844
Conc: 512.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Ankita
1/28/2021 2:54:24 PM



#30 AR-1254-5

R.T.: 7.343 min
Delta R.T.: -0.012 min
Response: 1948525
Conc: 430.52 ng/ml