

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012622\
 Data File : P0084444.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Jan 2022 21:58
 Operator : AJ\MA
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO012622

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 06:50:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012622.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jan 27 06:48:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	6.517	5.301	2834865	925884	49.458	48.445
2) SA Decachlor...	13.739	11.296	4102552	1430443	51.063	50.382
Target Compounds						
16) L4 AR-1242-1	7.967	6.768	870018	306275	493.058	486.179
17) L4 AR-1242-2	7.993	6.791	1268676	448575	494.157	497.612
18) L4 AR-1242-3	8.066	7.008	851431	242515	502.736	492.597
19) L4 AR-1242-4	8.181	7.111	636745	243030	486.015	495.410
20) L4 AR-1242-5	9.007	7.728	746872	317228	499.377	492.589

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012622\
Data File : P0084444.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Jan 2022 21:58
Operator : AJ\MA
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO012622

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 06:50:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012622.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jan 27 06:48:05 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

