

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012624\
 Data File : P0101357.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jan 2024 15:44
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 25 21:45:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 24 04:51:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.333	3.631	205.8E6	131.4E6	53.625	50.832
2) SA Decachlor...	10.059	8.676	115.3E6	89110336	49.501	47.024
Target Compounds						
3) L1 AR-1016-1	5.500	4.724	55623506	40156358	569.018	543.188
4) L1 AR-1016-2	5.523	4.743	79182286	55458928	575.637	562.677
5) L1 AR-1016-3	5.584	4.920	51731451	29722491	576.660	541.774
6) L1 AR-1016-4	5.683	4.962	40728709	25508576	569.854	501.956
7) L1 AR-1016-5	5.978	5.178	40320573	33234053	510.092	509.054
31) L7 AR-1260-1	7.108	6.218	69084111	59736453	484.417	484.110
32) L7 AR-1260-2	7.365	6.407	76495984	69068741	522.216	505.142
33) L7 AR-1260-3	7.726	6.562	57249511	62865165	490.291	490.045
34) L7 AR-1260-4	7.953	7.036	61453231	50376535	521.991	487.437
35) L7 AR-1260-5	8.269	7.278	116.2E6	107.5E6	530.074	512.143



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012624\
Data File : PO101357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jan 2024 15:44
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 25 21:45:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 24 04:51:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.333 min
Delta R.T.: -0.001 min
Response: 205842590
Conc: 53.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.631 min
Delta R.T.: 0.000 min
Response: 131367277
Conc: 50.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.059 min
Delta R.T.: 0.002 min
Response: 115344273
Conc: 49.50 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.676 min
Delta R.T.: 0.000 min
Response: 89110336
Conc: 47.02 ng/ml



#3 AR-1016-1


R.T.: 5.500 min
Delta R.T.: 0.000 min
Response: 55623506
Conc: 569.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.724 min
Delta R.T.: 0.000 min
Response: 40156358
Conc: 543.19 ng/ml




#4 AR-1016-2

R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 79182286
Conc: 575.64 ng/ml



#4 AR-1016-2


R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 55458928
Conc: 562.68 ng/ml



#5 AR-1016-3


R.T.: 5.584 min
Delta R.T.: 0.000 min
Response: 51731451
Conc: 576.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.920 min
Delta R.T.: 0.000 min
Response: 29722491
Conc: 541.77 ng/ml




#6 AR-1016-4

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 40728709
Conc: 569.85 ng/ml



#6 AR-1016-4


R.T.: 4.962 min
Delta R.T.: 0.000 min
Response: 25508576
Conc: 501.96 ng/ml



#7 AR-1016-5


R.T.: 5.978 min
Delta R.T.: 0.000 min
Response: 40320573
Conc: 510.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 33234053
Conc: 509.05 ng/ml




#31 AR-1260-1

R.T.: 7.108 min
Delta R.T.: 0.000 min
Response: 69084111
Conc: 484.42 ng/ml



#31 AR-1260-1


R.T.: 6.218 min
Delta R.T.: 0.000 min
Response: 59736453
Conc: 484.11 ng/ml



#32 AR-1260-2


R.T.: 7.365 min
Delta R.T.: 0.000 min
Response: 76495984
Conc: 522.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.407 min
Delta R.T.: 0.000 min
Response: 69068741
Conc: 505.14 ng/ml




#33 AR-1260-3

R.T.: 7.726 min
Delta R.T.: 0.000 min
Response: 57249511
Conc: 490.29 ng/ml



#33 AR-1260-3


R.T.: 6.562 min
Delta R.T.: 0.000 min
Response: 62865165
Conc: 490.05 ng/ml



#34 AR-1260-4


R.T.: 7.953 min
Delta R.T.: 0.000 min
Response: 61453231
Conc: 521.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.036 min
Delta R.T.: 0.000 min
Response: 50376535
Conc: 487.44 ng/ml



#35 AR-1260-5

R.T.: 8.269 min
Delta R.T.: 0.001 min
Response: 116239250
Conc: 530.07 ng/ml



#35 AR-1260-5

R.T.: 7.278 min
Delta R.T.: 0.000 min
Response: 107503553
Conc: 512.14 ng/ml