

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0012725\
 Data File : P0109119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2025 14:16
 Operator : YP/AJ
 Sample : Q1187-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 SP-COMP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 23:25:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0012125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 22 03:46:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.700	3.697	186.8E6	125.7E6	24.719	23.442
2)	SA Decachlor...	8.763	8.714	148.0E6	69823864	21.359	20.305
Target Compounds							
3)	L1 AR-1016-1	4.774	4.818	389784	704556	1.545	4.351 #
4)	L1 AR-1016-2	4.832	4.818	765500	704556	2.220	2.957 #
5)	L1 AR-1016-3	4.865	4.983	275019	387479	1.127	2.972 #
6)	L1 AR-1016-4	5.001	5.037	394794	199430	2.069	1.804
7)	L1 AR-1016-5	5.246	5.258	2673264	736506	12.805	5.131 #
8)	L2 AR-1221-1	3.904	3.921	541921	1242296	5.414	18.969 #
9)	L2 AR-1221-2	4.000	3.994	6298578	3941191	84.650	79.801
10)	L2 AR-1221-3	4.065	4.055	2100554	1052503	10.098	7.251 #
11)	L3 AR-1232-1	4.065	4.055	2100554	1052503	12.721	9.112 #
12)	L3 AR-1232-2	4.574	4.818	359950	704556	3.956	6.469 #
13)	L3 AR-1232-3	4.832	4.983	765500	387479	4.780	6.462 #
14)	L3 AR-1232-4	5.001	5.060	394794	883080	4.486	15.422 #
15)	L3 AR-1232-5	5.035	5.230	353033	2447948	5.447	40.347 #
16)	L4 AR-1242-1	4.774	4.775	389784	201395	1.835	1.409
17)	L4 AR-1242-2	4.832	4.818	765500	704556	2.670	3.572 #
18)	L4 AR-1242-3	4.865	4.983	275019	387479	1.343	3.553 #
19)	L4 AR-1242-4	5.001	5.060	394794	883080	2.485	7.771 #
20)	L4 AR-1242-5	5.652	5.587	866788	2702471	5.189	19.757 #
21)	L5 AR-1248-1	4.774	4.775	389784	201395	2.443	1.877
22)	L5 AR-1248-2	5.035	5.018	353033	536890	1.580	3.487 #
23)	L5 AR-1248-3	5.246	5.060	2673264	883080	9.761	5.340 #
24)	L5 AR-1248-4	5.606	5.230	3161286	2447948	8.221	12.783 #
25)	L5 AR-1248-5	5.652	5.629	866788	1222494	3.263	6.673 #
26)	L6 AR-1254-1	5.606	5.587	3161286	2702471	7.645	9.614 #
27)	L6 AR-1254-2	5.754	5.734	1089437	1925463	3.011	7.694 #
28)	L6 AR-1254-3	6.161	6.136	2852353	2196447	5.056	5.591
29)	L6 AR-1254-4	6.388	6.361	2374131	982663	6.889	4.460 #
30)	L6 AR-1254-5	6.809	6.779	9997219	6122811	19.484	18.741
31)	L7 AR-1260-1	6.293	6.293	4333623	258529	11.370	1.026 #
32)	L7 AR-1260-2	6.481	6.492	3421142	290043	7.286	0.964 #
33)	L7 AR-1260-3	6.849	6.608	941653	1441807	2.405	5.179 #
34)	L7 AR-1260-4	7.105	7.113	1928654	749197	5.388	3.320 #
35)	L7 AR-1260-5	7.348	7.358	2310219	59253	2.768	0.118 #
36)	L8 AR-1262-1	6.849	6.819	941653	378444	1.763	1.112 #
37)	L8 AR-1262-2	7.348	7.317	2310219	1444345	2.513	2.604
38)	L8 AR-1262-3	7.639	7.602	976517	343351	2.679	1.601 #
39)	L8 AR-1262-4	7.699	7.663	3454764	1473253	5.024	3.759 #
40)	L8 AR-1262-5	8.177	8.135	3725004	231693	12.330	1.453 #
41)	L9 AR-1268-1	7.639	7.602	976517	343351	0.917	0.545 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012725\
 Data File : P0109119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2025 14:16
 Operator : YP/AJ
 Sample : Q1187-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SP-COMP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 23:25:29 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 22 03:46:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	7.699	7.663	3454764	1473253	3.565	2.580 #
43) L9	AR-1268-3	7.909	7.876	586709	413265	0.728	0.897
44) L9	AR-1268-4	8.177	8.135	3725004	231693	10.874	1.284 #
45) L9	AR-1268-5	8.499	8.453	1843109	953063	0.771	0.796



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012725\
Data File : PO109119.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2025 14:16
Operator : YP/AJ
Sample : Q1187-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SP-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 23:25:29 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 22 03:46:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.700 min
Delta R.T.: 0.000 min
Response: 186789591
Conc: 24.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#1 Tetrachloro-m-xylene

R.T.: 3.697 min
Delta R.T.: -0.022 min
Response: 125657098
Conc: 23.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.763 min
Delta R.T.: 0.005 min
Response: 147967543
Conc: 21.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.714 min
Delta R.T.: -0.019 min
Response: 69823864
Conc: 20.31 ng/ml



#3 AR-1016-1


R.T.: 4.774 min
Delta R.T.: -0.021 min
Response: 389784
Conc: 1.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#3 AR-1016-1

R.T.: 4.818 min
Delta R.T.: 0.015 min
Response: 704556
Conc: 4.35 ng/ml




#4 AR-1016-2

R.T.: 4.832 min
Delta R.T.: 0.018 min
Response: 765500
Conc: 2.22 ng/ml



#4 AR-1016-2


R.T.: 4.818 min
Delta R.T.: -0.004 min
Response: 704556
Conc: 2.96 ng/ml



#5 AR-1016-3


R.T.: 4.865 min
Delta R.T.: -0.005 min
Response: 275019
Conc: 1.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#5 AR-1016-3

R.T.: 4.983 min
Delta R.T.: -0.015 min
Response: 387479
Conc: 2.97 ng/ml




#6 AR-1016-4

R.T.: 5.001 min
Delta R.T.: 0.010 min
Response: 394794
Conc: 2.07 ng/ml



#6 AR-1016-4


R.T.: 5.037 min
Delta R.T.: -0.003 min
Response: 199430
Conc: 1.80 ng/ml



#7 AR-1016-5


R.T.: 5.246 min
Delta R.T.: -0.004 min
Response: 2673264
Conc: 12.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#7 AR-1016-5

R.T.: 5.258 min
Delta R.T.: 0.006 min
Response: 736506
Conc: 5.13 ng/ml




#8 AR-1221-1

R.T.: 3.904 min
Delta R.T.: -0.010 min
Response: 541921
Conc: 5.41 ng/ml



#8 AR-1221-1


R.T.: 3.921 min
Delta R.T.: 0.012 min
Response: 1242296
Conc: 18.97 ng/ml



#9 AR-1221-2


R.T.: 4.000 min
Delta R.T.: 0.000 min
Response: 6298578
Conc: 84.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#9 AR-1221-2

R.T.: 3.994 min
Delta R.T.: 0.000 min
Response: 3941191
Conc: 79.80 ng/ml




#10 AR-1221-3

R.T.: 4.065 min
Delta R.T.: -0.012 min
Response: 2100554
Conc: 10.10 ng/ml



#10 AR-1221-3


R.T.: 4.055 min
Delta R.T.: -0.016 min
Response: 1052503
Conc: 7.25 ng/ml



#11 AR-1232-1


R.T.: 4.065 min
Delta R.T.: -0.012 min
Response: 2100554
Conc: 12.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#11 AR-1232-1

R.T.: 4.055 min
Delta R.T.: -0.016 min
Response: 1052503
Conc: 9.11 ng/ml




#12 AR-1232-2

R.T.: 4.574 min
Delta R.T.: 0.002 min
Response: 359950
Conc: 3.96 ng/ml



#12 AR-1232-2


R.T.: 4.818 min
Delta R.T.: 0.017 min
Response: 704556
Conc: 6.47 ng/ml



#13 AR-1232-3


R.T.: 4.832 min
Delta R.T.: 0.018 min
Response: 765500
Conc: 4.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#13 AR-1232-3

R.T.: 4.983 min
Delta R.T.: 0.007 min
Response: 387479
Conc: 6.46 ng/ml




#14 AR-1232-4

R.T.: 5.001 min
Delta R.T.: 0.009 min
Response: 394794
Conc: 4.49 ng/ml



#14 AR-1232-4


R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 883080
Conc: 15.42 ng/ml



#15 AR-1232-5


R.T.: 5.035 min
Delta R.T.: 0.002 min
Response: 353033
Conc: 5.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#15 AR-1232-5

R.T.: 5.230 min
Delta R.T.: -0.002 min
Response: 2447948
Conc: 40.35 ng/ml




#16 AR-1242-1

R.T.: 4.774 min
Delta R.T.: -0.021 min
Response: 389784
Conc: 1.83 ng/ml



#16 AR-1242-1


R.T.: 4.775 min
Delta R.T.: -0.007 min
Response: 201395
Conc: 1.41 ng/ml



#17 AR-1242-2


R.T.: 4.832 min
Delta R.T.: 0.017 min
Response: 765500
Conc: 2.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#17 AR-1242-2

R.T.: 4.818 min
Delta R.T.: 0.017 min
Response: 704556
Conc: 3.57 ng/ml




#18 AR-1242-3

R.T.: 4.865 min
Delta R.T.: -0.005 min
Response: 275019
Conc: 1.34 ng/ml



#18 AR-1242-3


R.T.: 4.983 min
Delta R.T.: 0.006 min
Response: 387479
Conc: 3.55 ng/ml



#19 AR-1242-4


R.T.: 5.001 min
Delta R.T.: 0.009 min
Response: 394794
Conc: 2.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#19 AR-1242-4

R.T.: 5.060 min
Delta R.T.: -0.002 min
Response: 883080
Conc: 7.77 ng/ml




#20 AR-1242-5

R.T.: 5.652 min
Delta R.T.: 0.006 min
Response: 866788
Conc: 5.19 ng/ml



#20 AR-1242-5


R.T.: 5.587 min
Delta R.T.: 0.004 min
Response: 2702471
Conc: 19.76 ng/ml



#21 AR-1248-1


R.T.: 4.774 min
Delta R.T.: -0.021 min
Response: 389784
Conc: 2.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#21 AR-1248-1

R.T.: 4.775 min
Delta R.T.: -0.006 min
Response: 201395
Conc: 1.88 ng/ml




#22 AR-1248-2

R.T.: 5.035 min
Delta R.T.: 0.002 min
Response: 353033
Conc: 1.58 ng/ml



#22 AR-1248-2


R.T.: 5.018 min
Delta R.T.: 0.000 min
Response: 536890
Conc: 3.49 ng/ml



#23 AR-1248-3


R.T.: 5.246 min
Delta R.T.: -0.003 min
Response: 2673264
Conc: 9.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#23 AR-1248-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 883080
Conc: 5.34 ng/ml




#24 AR-1248-4

R.T.: 5.606 min
Delta R.T.: 0.003 min
Response: 3161286
Conc: 8.22 ng/ml



#24 AR-1248-4


R.T.: 5.230 min
Delta R.T.: -0.002 min
Response: 2447948
Conc: 12.78 ng/ml



#25 AR-1248-5


R.T.: 5.652 min
Delta R.T.: 0.007 min
Response: 866788
Conc: 3.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#25 AR-1248-5

R.T.: 5.629 min
Delta R.T.: 0.006 min
Response: 1222494
Conc: 6.67 ng/ml




#26 AR-1254-1

R.T.: 5.606 min
Delta R.T.: 0.002 min
Response: 3161286
Conc: 7.64 ng/ml



#26 AR-1254-1


R.T.: 5.587 min
Delta R.T.: 0.003 min
Response: 2702471
Conc: 9.61 ng/ml



#27 AR-1254-2


R.T.: 5.754 min
Delta R.T.: 0.001 min
Response: 1089437
Conc: 3.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#27 AR-1254-2

R.T.: 5.734 min
Delta R.T.: 0.002 min
Response: 1925463
Conc: 7.69 ng/ml




#28 AR-1254-3

R.T.: 6.161 min
Delta R.T.: 0.000 min
Response: 2852353
Conc: 5.06 ng/ml



#28 AR-1254-3


R.T.: 6.136 min
Delta R.T.: 0.002 min
Response: 2196447
Conc: 5.59 ng/ml



#29 AR-1254-4


R.T.: 6.388 min
Delta R.T.: 0.000 min
Response: 2374131
Conc: 6.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#29 AR-1254-4

R.T.: 6.361 min
Delta R.T.: 0.000 min
Response: 982663
Conc: 4.46 ng/ml




#30 AR-1254-5

R.T.: 6.809 min
Delta R.T.: 0.000 min
Response: 9997219
Conc: 19.48 ng/ml



#30 AR-1254-5


R.T.: 6.779 min
Delta R.T.: 0.000 min
Response: 6122811
Conc: 18.74 ng/ml



#31 AR-1260-1


R.T.: 6.293 min
Delta R.T.: 0.003 min
Response: 4333623
Conc: 11.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#31 AR-1260-1

R.T.: 6.293 min
Delta R.T.: 0.006 min
Response: 258529
Conc: 1.03 ng/ml




#32 AR-1260-2

R.T.: 6.481 min
Delta R.T.: 0.003 min
Response: 3421142
Conc: 7.29 ng/ml



#32 AR-1260-2


R.T.: 6.492 min
Delta R.T.: 0.019 min
Response: 290043
Conc: 0.96 ng/ml



#33 AR-1260-3


R.T.: 6.849 min
Delta R.T.: 0.002 min
Response: 941653
Conc: 2.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#33 AR-1260-3

R.T.: 6.608 min
Delta R.T.: -0.019 min
Response: 1441807
Conc: 5.18 ng/ml




#34 AR-1260-4

R.T.: 7.105 min
Delta R.T.: -0.003 min
Response: 1928654
Conc: 5.39 ng/ml



#34 AR-1260-4


R.T.: 7.113 min
Delta R.T.: 0.014 min
Response: 749197
Conc: 3.32 ng/ml



#35 AR-1260-5


R.T.: 7.348 min
Delta R.T.: -0.001 min
Response: 2310219
Conc: 2.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#35 AR-1260-5

R.T.: 7.358 min
Delta R.T.: 0.018 min
Response: 59253
Conc: 0.12 ng/ml




#36 AR-1262-1

R.T.: 6.849 min
Delta R.T.: 0.001 min
Response: 941653
Conc: 1.76 ng/ml



#36 AR-1262-1


R.T.: 6.819 min
Delta R.T.: 0.000 min
Response: 378444
Conc: 1.11 ng/ml



#37 AR-1262-2


R.T.: 7.348 min
Delta R.T.: 0.000 min
Response: 2310219
Conc: 2.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#37 AR-1262-2

R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 1444345
Conc: 2.60 ng/ml




#38 AR-1262-3

R.T.: 7.639 min
Delta R.T.: 0.004 min
Response: 976517
Conc: 2.68 ng/ml



#38 AR-1262-3


R.T.: 7.602 min
Delta R.T.: 0.000 min
Response: 343351
Conc: 1.60 ng/ml



#39 AR-1262-4


R.T.: 7.699 min
Delta R.T.: 0.000 min
Response: 3454764
Conc: 5.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#39 AR-1262-4

R.T.: 7.663 min
Delta R.T.: -0.002 min
Response: 1473253
Conc: 3.76 ng/ml




#40 AR-1262-5

R.T.: 8.177 min
Delta R.T.: -0.019 min
Response: 3725004
Conc: 12.33 ng/ml



#40 AR-1262-5


R.T.: 8.135 min
Delta R.T.: -0.023 min
Response: 231693
Conc: 1.45 ng/ml



#41 AR-1268-1


R.T.: 7.639 min
Delta R.T.: 0.003 min
Response: 976517
Conc: 0.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#41 AR-1268-1

R.T.: 7.602 min
Delta R.T.: 0.001 min
Response: 343351
Conc: 0.55 ng/ml




#42 AR-1268-2

R.T.: 7.699 min
Delta R.T.: 0.000 min
Response: 3454764
Conc: 3.56 ng/ml



#42 AR-1268-2


R.T.: 7.663 min
Delta R.T.: -0.002 min
Response: 1473253
Conc: 2.58 ng/ml



#43 AR-1268-3


R.T.: 7.909 min
Delta R.T.: 0.000 min
Response: 586709
Conc: 0.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP




#43 AR-1268-3

R.T.: 7.876 min
Delta R.T.: 0.003 min
Response: 413265
Conc: 0.90 ng/ml




#44 AR-1268-4

R.T.: 8.177 min
Delta R.T.: -0.019 min
Response: 3725004
Conc: 10.87 ng/ml



#44 AR-1268-4


R.T.: 8.135 min
Delta R.T.: -0.022 min
Response: 231693
Conc: 1.28 ng/ml



#45 AR-1268-5

R.T.: 8.499 min
Delta R.T.: 0.004 min
Response: 1843109
Conc: 0.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-COMP



#45 AR-1268-5

R.T.: 8.453 min
Delta R.T.: 0.000 min
Response: 953063
Conc: 0.80 ng/ml